

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090418\
 Data File : VX004345.D
 Acq On : 04 Sep 2018 14:25
 Operator : JC/MD
 Sample : J4762-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE BLANK-WATERREF-4

Quant Time: Sep 05 07:09:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	276757	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	439720	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	415410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	224553	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193988	53.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.56%	
35) Dibromofluoromethane	5.50	113	170493	52.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.24%	
50) Toluene-d8	8.71	98	628443	50.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.14	95	222455	48.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090418\
Data File : VX004345.D
Acq On : 04 Sep 2018 14:25
Operator : JC/MD
Sample : J4762-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-WATERREF-4

Quant Time: Sep 05 07:09:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

