

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090418\
 Data File : VX004353.D
 Acq On : 04 Sep 2018 17:59
 Operator : JC/MD
 Sample : J4747-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-DRUM-3-STONE

Quant Time: Sep 05 07:36:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	267829	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	435072	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	422281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	230578	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188902	54.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.22%	
35) Dibromofluoromethane	5.50	113	165485	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	8.71	98	626327	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	11.14	95	223077	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
Target Compounds						
18) Methyl Acetate	2.78	43	6910	1.33	ug/l	98
20) Methylene Chloride	2.85	84	14993	4.46	ug/l	95
25) 2-Butanone	4.71	43	10110	1.49	ug/l	99
30) Chloroform	5.21	83	6777	1.19	ug/l	91
43) Isopropyl Acetate	6.46	43	147116	19.90	ug/l	100
95) Naphthalene	13.83	128	66612	4.43	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090418\
Data File : VX004353.D
Acq On : 04 Sep 2018 17:59
Operator : JC/MD
Sample : J4747-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-DRUM-3-STONE

Quant Time: Sep 05 07:36:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

