

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090418\
 Data File : VX004357.D
 Acq On : 04 Sep 2018 19:47
 Operator : JC/MD
 Sample : J4693-23
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-MW-EB-01

Quant Time: Sep 05 08:04:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	297274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	463530	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	426039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	241781	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	196907	50.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.64%	
35) Dibromofluoromethane	5.50	113	173217	50.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.46%	
50) Toluene-d8	8.71	98	651321	50.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.14	95	230988	47.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090418\
Data File : VX004357.D
Acq On : 04 Sep 2018 19:47
Operator : JC/MD
Sample : J4693-23
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-MW-EB-01

Quant Time: Sep 05 08:04:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

