

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090419\
 Data File : VX012070.D
 Acq On : 05 Sep 2019 00:33
 Operator : SY/SP
 Sample : VX0904SBSD02
 Misc : 5.00g/5.0mL/MSVOA_X/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02

Quant Time: Sep 05 05:28:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	117085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	230968	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	215455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103863	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.04	65	91830	56.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.68%	
35) Dibromofluoromethane	5.48	113	72127	50.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.56%	
50) Toluene-d8	8.71	98	266387	51.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.13	95	117876	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.19	85	19655	18.204	ug/l	91
3) Chloromethane	1.31	50	21469	19.415	ug/l	99
4) Vinyl Chloride	1.39	62	20235	19.450	ug/l	94
5) Bromomethane	1.62	94	10202	19.768	ug/l	97
6) Chloroethane	1.70	64	12851	19.870	ug/l	98
7) Trichlorofluoromethane	1.91	101	29497	18.255	ug/l	100
8) Diethyl Ether	2.17	74	12248	21.934	ug/l	92
9) 1,1,2-Trichlorotrifluoroet	2.37	101	22436	18.746	ug/l	# 83
10) Methyl Iodide	2.50	142	14885	17.556	ug/l	92
11) Tert butyl alcohol	3.04	59	17841	114.975	ug/l	96
12) 1,1-Dichloroethene	2.36	96	21364	18.530	ug/l	92
13) Acrolein	2.28	56	5395	121.527	ug/l	94
14) Allyl chloride	2.71	41	42128	20.194	ug/l	99
15) Acrylonitrile	3.13	53	38886	110.822	ug/l	98
16) Acetone	2.43	43	35327	91.485	ug/l	92
17) Carbon Disulfide	2.55	76	55804	16.562	ug/l	98
18) Methyl Acetate	2.76	43	24190	26.592	ug/l	99
19) Methyl tert-butyl Ether	3.18	73	83961	22.369	ug/l	99
20) Methylene Chloride	2.84	84	32043	20.664	ug/l	95
21) trans-1,2-Dichloroethene	3.15	96	26096	19.288	ug/l	89
22) Diisopropyl ether	3.84	45	94674	22.265	ug/l	98
23) Vinyl Acetate	3.80	43	276121	100.383	ug/l	99
24) 1,1-Dichloroethane	3.68	63	52233	21.180	ug/l	99
25) 2-Butanone	4.66	43	53054	107.935	ug/l	99
26) 2,2-Dichloropropane	4.56	77	38894	17.082	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	32968	20.788	ug/l	93
28) Bromochloromethane	5.00	49	23261	23.629	ug/l	# 80
29) Tetrahydrofuran	5.12	42	32789	110.717	ug/l	98
30) Chloroform	5.19	83	57049	21.559	ug/l	97
31) Cyclohexane	5.56	56	35424	17.885	ug/l	97
32) 1,1,1-Trichloroethane	5.48	97	47054	20.416	ug/l	97
36) 1,1-Dichloropropene	5.79	75	35375	17.390	ug/l	94
37) Ethyl Acetate	4.83	43	22656	20.481	ug/l	98
38) Carbon Tetrachloride	5.77	117	37929	17.694	ug/l	96

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090419\
 Data File : VX012070.D
 Acq On : 05 Sep 2019 00:33
 Operator : SY/SP
 Sample : VX0904SBSD02
 Misc : 5.00g/5.0mL/MSVOA_X/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VX0904SBSD02

Quant Time: Sep 05 05:28:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.45	83	38441	16.282	ug/l	97
40) Benzene	6.13	78	110675	18.951	ug/l	99
41) Methacrylonitrile	5.03	41	13622	20.649	ug/l	94
42) 1,2-Dichloroethane	6.18	62	44643	21.243	ug/l	97
43) Isopropyl Acetate	6.44	43	46606	20.720	ug/l	99
44) Trichloroethene	7.20	130	29420	18.678	ug/l	81
45) 1,2-Dichloropropane	7.51	63	30407	20.153	ug/l	100
46) Dibromomethane	7.65	93	17585	19.965	ug/l #	75
47) Bromodichloromethane	7.89	83	44711	19.970	ug/l	95
48) Methyl methacrylate	7.76	41	23487	20.767	ug/l	95
49) 1,4-Dioxane	7.73	88	6594	427.699	ug/l	95
51) 4-Methyl-2-Pentanone	8.64	43	120611	105.423	ug/l	99
52) Toluene	8.78	92	71631	18.852	ug/l	99
53) t-1,3-Dichloropropene	9.04	75	42850	18.487	ug/l	99
54) cis-1,3-Dichloropropene	8.43	75	48675	18.884	ug/l	99
55) 1,1,2-Trichloroethane	9.21	97	26062	20.696	ug/l	98
56) Ethyl methacrylate	9.17	69	36600	20.291	ug/l	97
57) 1,3-Dichloropropane	9.37	76	44580	20.516	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.30	63	85031	104.566	ug/l	94
59) 2-Hexanone	9.49	43	85350	102.240	ug/l	97
60) Dibromochloromethane	9.57	129	30400	20.256	ug/l	99
61) 1,2-Dibromoethane	9.67	107	24468	20.109	ug/l	98
64) Tetrachloroethene	9.33	164	27840	20.726	ug/l #	89
65) Chlorobenzene	10.14	112	81756	19.232	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.21	131	30563	19.662	ug/l	97
67) Ethyl Benzene	10.25	91	144553	18.655	ug/l	96
68) m/p-Xylenes	10.36	106	105293	36.880	ug/l	89
69) o-Xylene	10.70	106	52075	18.810	ug/l	88
70) Styrene	10.71	104	92567	19.250	ug/l	95
71) Bromoform	10.86	173	16618	19.465	ug/l #	100
73) Isopropylbenzene	11.01	105	139687	18.330	ug/l	97
74) N-amyl acetate	10.89	43	43172	19.161	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.26	83	32109	20.211	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	35213	27.748	ug/l	76
77) Bromobenzene	11.25	156	34392	20.179	ug/l	69
78) n-propylbenzene	11.35	91	164266	17.842	ug/l	95
79) 2-Chlorotoluene	11.42	91	104550	18.745	ug/l	91
80) 1,3,5-Trimethylbenzene	11.50	105	125846	18.844	ug/l	94
81) trans-1,4-Dichloro-2-buten	11.07	75	8806	16.405	ug/l	91
82) 4-Chlorotoluene	11.51	91	129551	19.487	ug/l	89
83) tert-Butylbenzene	11.76	119	118096	18.553	ug/l	88
84) 1,2,4-Trimethylbenzene	11.80	105	127845	18.626	ug/l	96
85) sec-Butylbenzene	11.94	105	143937	18.206	ug/l	93
86) p-Isopropyltoluene	12.06	119	130008	18.082	ug/l	94
87) 1,3-Dichlorobenzene	12.02	146	66184	19.379	ug/l	93
88) 1,4-Dichlorobenzene	12.09	146	66837	19.573	ug/l	94
89) n-Butylbenzene	12.39	91	124488	17.824	ug/l	96
90) Hexachloroethane	12.59	117	26291	18.313	ug/l	57
91) 1,2-Dichlorobenzene	12.39	146	61624	20.035	ug/l	93
92) 1,2-Dibromo-3-Chloropropan	12.99	75	6596	19.841	ug/l	66

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090419\
Data File : VX012070.D
Acq On : 05 Sep 2019 00:33
Operator : SY/SP
Sample : VX0904SBSD02
Misc : 5.00g/5.0mL/MSVOA_X/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02

Quant Time: Sep 05 05:28:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.64	180	39866	18.210	ug/l	96
94) Hexachlorobutadiene	13.78	225	19864	18.000	ug/l	98
95) Naphthalene	13.83	128	92273	19.474	ug/l	99
96) 1,2,3-Trichlorobenzene	14.01	180	36109	18.748	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090419\
Data File : VX012070.D
Acq On : 05 Sep 2019 00:33
Operator : SY/SP
Sample : VX0904SBSD02
Misc : 5.00g/5.0mL/MSVOA_X/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02

Quant Time: Sep 05 05:28:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.64 min Scan# 780

Delta R.T. -0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBSD02

Tgt Ion:168 Resp: 117085

Ion Ratio Lower Upper

168 100

99 71.7 37.6 56.4#




Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0


5.64

Time-->



5.64

Time-->



#2

Dichlorodifluoromethane

Concen: 18.204 ug/l

RT: 1.19 min Scan# 49

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 85 Resp: 19655

Ion Ratio Lower Upper

85 100

87 36.4 15.6 46.8




Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.19

Time-->





#3

Chloromethane

Concen: 19.415 ug/l

RT: 1.31 min Scan# 70

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBS02

Tgt Ion: 50 Resp: 21469

Ion Ratio Lower Upper




50 100

52 33.5 26.5 39.7



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0



#4

Vinyl Chloride

Concen: 19.450 ug/l

RT: 1.39 min Scan# 83

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 62 Resp: 20235

Ion Ratio Lower Upper



62 100


64 29.3 26.2 39.4



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0





#5

Bromomethane

Concen: 19.768 ug/l

RT: 1.62 min Scan# 121

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBSD02

Tgt Ion: 94 Resp: 10202

Ion Ratio Lower Upper

94 100

96 92.9 76.3 114.5



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

1.62

12000

10000

8000

6000



4000

2000

0

1.60 1.65 1.70

Time-->



#6

Chloroethane

Concen: 19.870 ug/l

RT: 1.70 min Scan# 134

Delta R.T. -0.01 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 64 Resp: 12851

Ion Ratio Lower Upper

64 100

66 31.7 26.3 39.5



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

1.70

10000

8000

6000


4000


2000

0

1.65 1.70 1.75 1.80 1.85

Time-->









#7

Trichlorofluoromethane
 Concen: 18.255 ug/l
 RT: 1.91 min Scan# 168
 Delta R.T. -0.01 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02




Tgt Ion:101 Resp: 29497
 Ion Ratio Lower Upper
 101 100
 103 64.6 51.6 77.4




#8

Diethyl Ether
 Concen: 21.934 ug/l
 RT: 2.17 min Scan# 211
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 74 Resp: 12248
 Ion Ratio Lower Upper
 74 100
 45 99.1 45.9 137.6





#9

1,1,2-Trichlorotrifluoroethane

Concen: 18.746 ug/l

RT: 2.37 min Scan# 243

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion:101 Resp: 22436

Ion Ratio Lower Upper

101 100

85 45.8 34.5 51.7




151 71.6 75.0 112.4#



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 17.556 ug/l

RT: 2.50 min Scan# 264

Delta R.T. 0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33


Tgt Ion:142 Resp: 14885

Ion Ratio Lower Upper

142 100

127 50.6 35.4 53.2



141 15.1 11.4 17.2




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11
Tert butyl alcohol
Concen: 114.975 ug/l
RT: 3.04 min Scan# 353
Delta R.T. 0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion: 59 Resp: 17841
Ion Ratio Lower Upper
59 100
57 11.4 7.9 11.9



#12
1,1-Dichloroethene
Concen: 18.530 ug/l
RT: 2.36 min Scan# 241
Delta R.T. -0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 96 Resp: 21364
Ion Ratio Lower Upper
96 100
61 167.9 124.1 186.1
98 64.4 50.2 75.2





#13

Acrolein

Concen: 121.527 ug/l

RT: 2.28 min Scan# 228

Delta R.T. 0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :





VX0904SBS02

Tgt Ion: 56 Resp: 5395

Ion Ratio Lower Upper

56 100

55 68.7 59.2 88.8



#14

Allyl chloride

Concen: 20.194 ug/l

RT: 2.71 min Scan# 299

Delta R.T. -0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33




Tgt Ion: 41 Resp: 42128


Ion Ratio Lower Upper

41 100

39 72.7 57.9 86.9

76 34.6 28.5 42.7





#15

Acrylonitrile

Concen: 110.822 ug/l

RT: 3.13 min Scan# 368

Delta R.T. 0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion: 53 Resp: 38886

Ion Ratio Lower Upper

53 100

52 80.5 65.6 98.4




51 37.9 29.8 44.6



Abundance Ion 53.00 (52.70 to 53.70): VX0

Ion 52.00 (51.70 to 52.70): VX0

Ion 51.00 (50.70 to 51.70): VX0



#16

Acetone

Concen: 91.485 ug/l

RT: 2.43 min Scan# 253

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 43 Resp: 35327

Ion Ratio Lower Upper



43 100


58 32.6 22.8 34.2



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#17

Carbon Disulfide

Concen: 16.562 ug/l

RT: 2.55 min Scan# 273

Delta R.T. -0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBSD02

Tgt Ion: 76 Resp: 55804

Ion Ratio Lower Upper




76 100

78 8.7 7.4 11.2



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#18

Methyl Acetate

Concen: 26.592 ug/l

RT: 2.76 min Scan# 307

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 43 Resp: 24190

Ion Ratio Lower Upper



43 100


74 24.8 19.5 29.3



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0









#19
Methyl tert-butyl Ether
Concen: 22.369 ug/l
RT: 3.18 min Scan# 376
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBS02


Tgt Ion: 73 Resp: 83961
Ion Ratio Lower Upper
73 100
57 21.2 16.6 25.0



#20
Methylene Chloride
Concen: 20.664 ug/l
RT: 2.84 min Scan# 320
Delta R.T. -0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 84 Resp: 32043
Ion Ratio Lower Upper
84 100
49 127.4 95.8 143.8
51 39.1 28.6 43.0
86 64.1 50.9 76.3





#21

trans-1,2-Dichloroethene

Concen: 19.288 ug/l

RT: 3.15 min Scan# 371

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion: 96 Resp: 26096

Ion Ratio Lower Upper

96 100

61 151.2 107.0 160.6

98 65.8 49.8 74.6






Abundance

Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 22.265 ug/l

RT: 3.84 min Scan# 484

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Tgt Ion: 45 Resp: 94674


Ion Ratio Lower Upper

45 100

43 70.1 54.2 81.4

87 28.7 23.4 35.0

59 12.3 9.0 13.6





Abundance


Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 100.383 ug/l

RT: 3.80 min Scan# 478

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBSD02

Tgt Ion: 43 Resp: 276121

Ion Ratio Lower Upper




43 100

86 10.7 9.0 13.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 21.180 ug/l

RT: 3.68 min Scan# 458

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33


Tgt Ion: 63 Resp: 52233

Ion Ratio Lower Upper

63 100

98 7.9 3.9 11.7



100 4.6 2.5 7.4




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0





#25

2-Butanone

Concen: 107.935 ug/l

RT: 4.66 min Scan# 619

Delta R.T. 0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBS02

Tgt Ion: 43 Resp: 53054

Ion Ratio Lower Upper




43 100

72 27.1 21.3 31.9



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0



#26

2,2-Dichloropropane

Concen: 17.082 ug/l

RT: 4.56 min Scan# 603

Delta R.T. -0.01 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 77 Resp: 38894

Ion Ratio Lower Upper



77 100


97 22.3 11.5 34.5



Abundance Ion 76.90 (76.60 to 77.60): VX0

Ion 96.90 (96.60 to 97.60): VX0






#27




cis-1,2-Dichloroethene
 Concen: 20.788 ug/l
 RT: 4.58 min Scan# 605
 Delta R.T. -0.01 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02

Tgt Ion: 96 Resp: 32968
 Ion Ratio Lower Upper
 96 100
 61 150.6 0.0 277.0
 98 65.1 0.0 128.4




Abundance Ion 95.90 (95.60 to 96.60): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 97.90 (97.60 to 98.60): VX0





#28


Bromochloromethane
 Concen: 23.629 ug/l
 RT: 5.00 min Scan# 675
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 49 Resp: 23261
 Ion Ratio Lower Upper
 49 100
 129 1.2 0.0 5.6
 130 75.7 76.7 115.1#



Abundance Ion 48.90 (48.60 to 49.60): VX0
 Ion 128.80 (128.50 to 129.50): V
 Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 110.717 ug/l
RT: 5.12 min Scan# 694
Delta R.T. 0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion: 42 Resp: 32789
Ion Ratio Lower Upper
42 100
72 45.4 37.0 55.4
71 43.6 33.8 50.8



#30
Chloroform
Concen: 21.559 ug/l
RT: 5.19 min Scan# 706
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 83 Resp: 57049
Ion Ratio Lower Upper
83 100
85 63.1 52.6 78.8









#31
Cyclohexane
Concen: 17.885 ug/l
RT: 5.56 min Scan# 767
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion: 56 Resp: 35424
Ion Ratio Lower Upper
56 100
69 35.6 25.7 38.5
84 89.0 70.3 105.5



#32
1,1,1-Trichloroethane
Concen: 20.416 ug/l
RT: 5.48 min Scan# 753
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 97 Resp: 47054
Ion Ratio Lower Upper
97 100
99 64.7 52.1 78.1
61 45.0 32.8 49.2









#33
 1,2-Dichloroethane-d4
 Concen: 56.842 ug/l
 RT: 6.04 min Scan# 846
 Delta R.T. -0.01 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02


Tgt Ion: 65 Resp: 91830
 Ion Ratio Lower Upper
 65 100
 67 49.5 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 114 Resp: 230968
 Ion Ratio Lower Upper
 114 100
 63 22.8 0.0 37.4
 88 17.4 0.0 29.6









#35
 Dibromofluoromethane
 Concen: 50.780 ug/l
 RT: 5.48 min Scan# 754
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02


Tgt Ion:113 Resp: 72127
 Ion Ratio Lower Upper
 113 100
 111 103.2 81.7 122.5
 192 17.4 22.5 33.7#



#36
 1,1-Dichloropropene
 Concen: 17.390 ug/l
 RT: 5.79 min Scan# 804
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 75 Resp: 35375
 Ion Ratio Lower Upper
 75 100
 110 33.9 19.5 58.5
 77 32.5 24.6 37.0





#37

Ethyl Acetate

Concen: 20.481 ug/l

RT: 4.83 min Scan# 646

Delta R.T. 0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion: 43 Resp: 22656

Ion Ratio Lower Upper

43 100

61 14.7 10.9 16.3




70 12.3 9.4 14.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 70.00 (69.70 to 70.70): VX0



#38

Carbon Tetrachloride

Concen: 17.694 ug/l

RT: 5.77 min Scan# 801

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33


Tgt Ion: 117 Resp: 37929

Ion Ratio Lower Upper

117 100

119 93.3 78.3 117.5



121 31.9 24.2 36.2




Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V









#39
Methylcyclohexane
Concen: 16.282 ug/l
RT: 7.45 min Scan# 1077
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion: 83 Resp: 38441
Ion Ratio Lower Upper
83 100
55 78.9 59.6 89.4
98 51.1 41.4 62.0



#40
Benzene
Concen: 18.951 ug/l
RT: 6.13 min Scan# 860
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 78 Resp: 110675
Ion Ratio Lower Upper
78 100
77 24.4 19.3 28.9









#41
Methacrylonitrile
Concen: 20.649 ug/l
RT: 5.03 min Scan# 679
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion: 41 Resp: 13622
Ion Ratio Lower Upper
41 100
39 54.2 46.7 70.1
67 75.4 57.0 85.6
52 34.6 23.5 35.3



#42
1,2-Dichloroethane
Concen: 21.243 ug/l
RT: 6.18 min Scan# 868
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 62 Resp: 44643
Ion Ratio Lower Upper
62 100
98 8.5 0.0 19.2





#43

Isopropyl Acetate

Concen: 20.720 ug/l

RT: 6.44 min Scan# 910

Delta R.T. 0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion: 43 Resp: 46606

Ion Ratio Lower Upper

43 100

61 19.6 16.1 24.1




87 13.6 11.3 16.9



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 18.678 ug/l

RT: 7.20 min Scan# 1036

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 130 Resp: 29420

Ion Ratio Lower Upper



130 100


95 105.5 0.0 176.4



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0









#45
 1,2-Dichloropropane
 Concen: 20.153 ug/l
 RT: 7.51 min Scan# 1086
 Delta R.T. 0.01 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02


Tgt Ion: 63 Resp: 30407
 Ion Ratio Lower Upper
 63 100
 65 31.4 25.0 37.6



#46
 Dibromomethane
 Concen: 19.965 ug/l
 RT: 7.65 min Scan# 1110
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 93 Resp: 17585
 Ion Ratio Lower Upper
 93 100
 95 83.1 66.2 99.2
 174 89.6 110.2 165.4#









#47
 Bromodichloromethane
 Concen: 19.970 ug/l
 RT: 7.89 min Scan# 1149
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02


Tgt Ion: 83 Resp: 44711
 Ion Ratio Lower Upper
 83 100
 85 66.8 50.3 75.5
 127 8.2 8.1 12.1



#48
 Methyl methacrylate
 Concen: 20.767 ug/l
 RT: 7.76 min Scan# 1128
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 41 Resp: 23487
 Ion Ratio Lower Upper
 41 100
 69 75.9 66.3 99.5
 39 61.6 48.1 72.1









#49
1,4-Dioxane
Concen: 427.699 ug/l
RT: 7.73 min Scan# 1123
Delta R.T. 0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02

Tgt Ion: 88 Resp: 6594
Ion Ratio Lower Upper
88 100
43 36.5 29.8 44.6
58 75.3 55.4 83.2




Abundance Ion 88.00 (87.70 to 88.70): VX0
Ion 43.00 (42.70 to 43.70): VX0
Ion 58.00 (57.70 to 58.70): VX0





#50
Toluene-d8
Concen: 51.109 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 98 Resp: 266387
Ion Ratio Lower Upper
98 100
100 66.4 53.3 79.9



Abundance Ion 98.00 (97.70 to 98.70): VX0
Ion 100.00 (99.70 to 100.70): VX0









#51
 4-Methyl-2-Pentanone
 Concen: 105.423 ug/l
 RT: 8.64 min Scan# 1271
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02


Tgt Ion: 43 Resp: 120611
 Ion Ratio Lower Upper
 43 100
 58 38.9 31.5 47.3



#52
 Toluene
 Concen: 18.852 ug/l
 RT: 8.78 min Scan# 1294
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 92 Resp: 71631
 Ion Ratio Lower Upper
 92 100
 91 171.2 136.0 204.0









#53
 t-1,3-Dichloropropene
 Concen: 18.487 ug/l
 RT: 9.04 min Scan# 1337
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02


Tgt Ion: 75 Resp: 42850
 Ion Ratio Lower Upper
 75 100
 77 31.1 25.1 37.7



#54
 cis-1,3-Dichloropropene
 Concen: 18.884 ug/l
 RT: 8.43 min Scan# 1237
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 75 Resp: 48675
 Ion Ratio Lower Upper
 75 100
 77 31.7 25.5 38.3
 39 54.6 43.0 64.4









#55
 1,1,2-Trichloroethane
 Concen: 20.696 ug/l
 RT: 9.21 min Scan# 1365
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02


Tgt Ion:	97	Resp:	26062
Ion	Ratio	Lower	Upper
97	100		
83	83.1	66.0	99.0
85	54.1	41.0	61.6
99	62.4	50.6	75.8



#56
 Ethyl methacrylate
 Concen: 20.291 ug/l
 RT: 9.17 min Scan# 1359
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion:	69	Resp:	36600
Ion	Ratio	Lower	Upper
69	100		
41	69.3	53.9	80.9
39	41.5	31.1	46.7









#57
 1,3-Dichloropropane
 Concen: 20.516 ug/l
 RT: 9.37 min Scan# 1391
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02


Tgt Ion: 76 Resp: 44580
 Ion Ratio Lower Upper
 76 100
 78 32.5 25.5 38.3



#58
 2-Chloroethyl Vinyl ether
 Concen: 104.566 ug/l
 RT: 8.30 min Scan# 1216
 Delta R.T. -0.01 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 63 Resp: 85031
 Ion Ratio Lower Upper
 63 100
 106 29.7 26.5 39.7









#59
2-Hexanone
Concen: 102.240 ug/l
RT: 9.49 min Scan# 1411
Delta R.T. 0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02

Tgt Ion: 43 Resp: 85350
Ion Ratio Lower Upper
43 100
58 52.1 27.2 81.6




Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 58.00 (57.70 to 58.70): VX0






#60
Dibromochloromethane
Concen: 20.256 ug/l
RT: 9.57 min Scan# 1425
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 129 Resp: 30400
Ion Ratio Lower Upper
129 100
127 77.3 38.1 114.5



Abundance Ion 128.80 (128.50 to 129.50): V
Ion 126.80 (126.50 to 127.50): V





#61

1,2-Dibromoethane

Concen: 20.109 ug/l

RT: 9.67 min Scan# 1440

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBSD02

Tgt Ion: 107 Resp: 24468

Ion Ratio Lower Upper

107 100

109 94.1 76.9 115.3



Abundance Ion 106.90 (106.60 to 107.60): V

20000 Ion 108.80 (108.50 to 109.50): V



15000

10000

5000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 48.629 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. -0.01 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33


Tgt Ion: 95 Resp: 117876

Ion Ratio Lower Upper

95 100

174 62.0 0.0 205.4

176 60.5 0.0 201.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

120000 Ion 173.80 (173.50 to 174.50): V

100000 Ion 175.80 (175.50 to 176.50): V

80000


60000


40000

20000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion:117 Resp: 215455
Ion Ratio Lower Upper
117 100
82 57.0 39.4 59.0
119 31.9 26.1 39.1



#64
Tetrachloroethene
Concen: 20.726 ug/l
RT: 9.33 min Scan# 1385
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion:164 Resp: 27840
Ion Ratio Lower Upper
164 100
166 130.5 104.3 156.5
129 106.5 68.3 102.5#
131 96.1 65.4 98.0









#65
Chlorobenzene
Concen: 19.232 ug/l
RT: 10.14 min Scan# 1517
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion:112 Resp: 81756
Ion Ratio Lower Upper
112 100
114 30.8 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 19.662 ug/l
RT: 10.21 min Scan# 1530
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion:131 Resp: 30563
Ion Ratio Lower Upper
131 100
133 94.9 48.4 145.0
119 64.6 30.5 91.5









#67
Ethyl Benzene
Concen: 18.655 ug/l
RT: 10.25 min Scan# 1535
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBS02

Tgt Ion: 91 Resp: 144553
Ion Ratio Lower Upper
91 100
106 30.0 25.8 38.8



#68
m/p-Xylenes
Concen: 36.880 ug/l
RT: 10.36 min Scan# 1553
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion: 106 Resp: 105293
Ion Ratio Lower Upper
106 100
91 213.0 157.5 236.3









#69
o-Xylene
Concen: 18.810 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion:106 Resp: 52075
Ion Ratio Lower Upper
106 100
91 225.7 103.5 310.3



#70
Styrene
Concen: 19.250 ug/l
RT: 10.71 min Scan# 1611
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion:104 Resp: 92567
Ion Ratio Lower Upper
104 100
78 56.7 39.1 58.7
103 54.3 43.4 65.2





#71

Bromoform

Concen: 19.465 ug/l

RT: 10.86 min Scan# 1635

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion:173 Resp: 16618

Ion Ratio Lower Upper




173 100

175 49.5 24.9 74.7

254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1835

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33


Tgt Ion:152 Resp: 103863

Ion Ratio Lower Upper



152 100


115 77.6 36.4 109.3

150 164.1 0.0 341.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 18.330 ug/l

RT: 11.01 min Scan# 1661

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBSD02

Tgt Ion:105 Resp: 139687

Ion Ratio Lower Upper




105 100

120 26.3 13.9 41.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 19.161 ug/l

RT: 10.89 min Scan# 1641

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Tgt Ion: 43 Resp: 43172


Ion Ratio Lower Upper

43 100

70 41.7 34.4 51.6

55 23.5 18.5 27.7

61 23.8 19.5 29.3






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#75

1,1,2,2-Tetrachloroethane

Concen: 20.211 ug/l

RT: 11.26 min Scan# 1702

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion: 83 Resp: 32109

Ion Ratio Lower Upper

83 100

131 10.3 6.3 19.1




85 64.5 32.5 97.4



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 27.748 ug/l

RT: 11.29 min Scan# 1707

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 75 Resp: 35213

Ion Ratio Lower Upper



75 100


77 30.9 23.5 70.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 20.179 ug/l

RT: 11.25 min Scan# 1700

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion:156 Resp: 34392

Ion Ratio Lower Upper

156 100

77 188.3 63.4 190.2




158 95.5 48.9 146.6



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 17.842 ug/l

RT: 11.35 min Scan# 1717

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 91 Resp: 164266

Ion Ratio Lower Upper



91 100


120 22.8 12.7 38.0



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 18.745 ug/l

RT: 11.42 min Scan# 1727

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :


VX0904SBS02

Tgt Ion: 91 Resp: 104550

Ion Ratio Lower Upper




91 100

126 31.9 18.6 55.8



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V



#80

1,3,5-Trimethylbenzene

Concen: 18.844 ug/l

RT: 11.50 min Scan# 1741

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 105 Resp: 125846

Ion Ratio Lower Upper



105 100


120 47.2 25.8 77.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 16.405 ug/l

RT: 11.07 min Scan# 1670

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02





Tgt Ion: 75 Resp: 8806

Ion Ratio Lower Upper

75 100

53 111.0 78.6 117.8

89 48.2 38.8 58.2



#82

4-Chlorotoluene

Concen: 19.487 ug/l

RT: 11.51 min Scan# 1742

Delta R.T. -0.00 min

Lab File: VX012070.D




Acq: 05 Sep 2019 00:33

Tgt Ion: 91 Resp: 129551

Ion Ratio Lower Upper

91 100

126 26.4 16.2 48.6









#83
tert-Butylbenzene
Concen: 18.553 ug/l
RT: 11.76 min Scan# 1784
Delta R.T. -0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33




Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02


Tgt Ion:119 Resp: 118096
Ion Ratio Lower Upper
119 100
91 64.9 27.1 81.2
134 21.9 12.1 36.3



#84
1,2,4-Trimethylbenzene
Concen: 18.626 ug/l
RT: 11.80 min Scan# 1790
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion:105 Resp: 127845
Ion Ratio Lower Upper
105 100
120 44.3 23.6 70.8





#85

sec-Butylbenzene

Concen: 18.206 ug/l

RT: 11.94 min Scan# 1813

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :





VX0904SBSD02

Tgt Ion:105 Resp: 143937

Ion Ratio Lower Upper

105 100

134 19.1 11.2 33.6



#86

p-Isopropyltoluene

Concen: 18.082 ug/l

RT: 12.06 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33




Tgt Ion:119 Resp: 130008


Ion Ratio Lower Upper

119 100

134 25.4 13.7 41.1

91 26.7 11.4 34.1









#87
 1,3-Dichlorobenzene
 Concen: 19.379 ug/l
 RT: 12.02 min Scan# 1826
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02

Tgt Ion:146 Resp: 66184
 Ion Ratio Lower Upper
 146 100
 111 44.2 18.1 54.1
 148 63.0 32.6 97.8




Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V






#88
 1,4-Dichlorobenzene
 Concen: 19.573 ug/l
 RT: 12.09 min Scan# 1838
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion:146 Resp: 66837
 Ion Ratio Lower Upper
 146 100
 111 45.2 17.7 53.1
 148 64.9 32.4 97.0



Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V





#89

n-Butylbenzene

Concen: 17.824 ug/l

RT: 12.39 min Scan# 1886

Delta R.T. -0.00 min

Lab File: VX012070.D

Acq: 05 Sep 2019 00:33

Instrument :

MSVOA_X

ClientSampleId :

VX0904SBSD02


Tgt Ion: 91 Resp: 124488

Ion Ratio Lower Upper

91 100

92 55.0 27.6 82.8




134 25.4 15.4 46.1



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 92.00 (91.70 to 92.70): VX0

Ion 134.00 (133.70 to 134.70): V



#90

Hexachloroethane

Concen: 18.313 ug/l

RT: 12.59 min Scan# 1920

Delta R.T. -0.00 min

Lab File: VX012070.D


Acq: 05 Sep 2019 00:33

Tgt Ion: 117 Resp: 26291

Ion Ratio Lower Upper



117 100


201 64.3 55.0 165.2



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V









#91
 1,2-Dichlorobenzene
 Concen: 20.035 ug/l
 RT: 12.39 min Scan# 1886
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBSD02


Tgt Ion:146 Resp: 61624
 Ion Ratio Lower Upper
 146 100
 111 47.6 18.6 56.0
 148 64.0 32.4 97.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 19.841 ug/l
 RT: 12.99 min Scan# 1986
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion: 75 Resp: 6596
 Ion Ratio Lower Upper
 75 100
 155 74.1 51.7 155.2
 157 89.4 66.9 200.7









#93
 1,2,4-Trichlorobenzene
 Concen: 18.210 ug/l
 RT: 13.64 min Scan# 2092
 Delta R.T. -0.00 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0904SBS02


Tgt Ion:180 Resp: 39866
 Ion Ratio Lower Upper
 180 100
 182 96.2 47.3 142.0
 145 34.9 13.7 41.0



#94
 Hexachlorobutadiene
 Concen: 18.000 ug/l
 RT: 13.78 min Scan# 2114
 Delta R.T. -0.01 min
 Lab File: VX012070.D
 Acq: 05 Sep 2019 00:33

Tgt Ion:225 Resp: 19864
 Ion Ratio Lower Upper
 225 100
 223 65.1 31.6 95.0
 227 63.2 32.5 97.5









#95
Naphthalene
Concen: 19.474 ug/l
RT: 13.83 min Scan# 2123
Delta R.T. -0.00 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Instrument :
MSVOA_X
ClientSampleId :
VX0904SBSD02

Tgt Ion:128 Resp: 92273
Ion Ratio Lower Upper
128 100
127 12.7 10.5 15.7
129 11.2 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 18.748 ug/l
RT: 14.01 min Scan# 2153
Delta R.T. -0.01 min
Lab File: VX012070.D
Acq: 05 Sep 2019 00:33

Tgt Ion:180 Resp: 36109
Ion Ratio Lower Upper
180 100
182 93.9 47.9 143.6
145 37.9 14.1 42.4

