

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090419\
 Data File : VX012050.D
 Acq On : 04 Sep 2019 14:24
 Operator : SY/SP
 Sample : K4546-14
 Misc : 5.75g/5.0mL/MSVOA X/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-S-B1-(2.5)

Quant Time: Sep 04 14:46:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	54788	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	90129	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	73389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	29610	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	32659	43.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.40%	
35) Dibromofluoromethane	5.49	113	26060	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	8.71	98	105349	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.14	95	39034	41.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.54%	
Target Compounds						
16) Acetone	2.43	43	3868	21.407	ug/l	95
20) Methylene Chloride	2.84	84	5359	7.386	ug/l	87
68) m/p-Xylenes	10.36	106	1380	1.419	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090419\
Data File : VX012050.D
Acq On : 04 Sep 2019 14:24
Operator : SY/SP
Sample : K4546-14
Misc : 5.75g/5.0mL/MSVOA X/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-S-B1-(2.5)

Quant Time: Sep 04 14:46:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

