

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090419\
 Data File : VX012063.D
 Acq On : 04 Sep 2019 20:26
 Operator : SY/SP
 Sample : K4545-01
 Misc : 4.73g/5.0mL/MSVOA X/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T16-17-E2-(0)

Quant Time: Sep 05 05:49:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	137373	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	238468	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	161720	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	33554	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	93459	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
35) Dibromofluoromethane	5.48	113	71826	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	8.71	98	253802	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
62) 4-Bromofluorobenzene	11.13	95	69148	27.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.26%	
Target Compounds						
11) Tert butyl alcohol	3.02	59	3731	20.493	ug/l #	77
16) Acetone	2.43	43	10478	23.127	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090419\
Data File : VX012063.D
Acq On : 04 Sep 2019 20:26
Operator : SY/SP
Sample : K4545-01
Misc : 4.73a/5.0mL/MSVOA X/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T16-17-E2-(0)

Quant Time: Sep 05 05:49:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

