

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090419\
 Data File : VX012072.D
 Acq On : 05 Sep 2019 01:29
 Operator : SY/SP
 Sample : K4545-04
 Misc : 5.34g/5.0mL/MSVOA X/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T16-17-E1-(2.25)

Quant Time: Sep 05 06:03:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	160902	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	283853	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	101606	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	108433	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
35) Dibromofluoromethane	5.48	113	84461	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	8.71	98	335503	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
62) 4-Bromofluorobenzene	11.13	95	127814	42.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.82%	
Target Compounds						
16) Acetone	2.43	43	14666	27.637	ug/l	97
20) Methylene Chloride	2.84	84	12429	5.833	ug/l	87
29) Tetrahydrofuran	5.12	42	3656	8.983	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090419\
Data File : VX012072.D
Acq On : 05 Sep 2019 01:29
Operator : SY/SP
Sample : K4545-04
Misc : 5.34g/5.0mL/MSVOA X/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T16-17-E1-(2.25)

Quant Time: Sep 05 06:03:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

