

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090419\
 Data File : VX012088.D
 Acq On : 05 Sep 2019 08:57
 Operator : SY/SP
 Sample : K4548-06
 Misc : 6.84g/5.0mL/MSVOA X/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T8-S-G2-(2.5)

Quant Time: Sep 05 14:51:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	152913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	264507	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	229811	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97838	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	103313	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
35) Dibromofluoromethane	5.48	113	77995	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
50) Toluene-d8	8.71	98	314261	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
62) 4-Bromofluorobenzene	11.13	95	122134	44.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.00%	
Target Compounds						
16) Acetone	2.43	43	16830	33.372	ug/l	98
20) Methylene Chloride	2.85	84	8833	4.362	ug/l	95
29) Tetrahydrofuran	5.13	42	3596	9.297	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090419\
Data File : VX012088.D
Acq On : 05 Sep 2019 08:57
Operator : SY/SP
Sample : K4548-06
Misc : 6.84g/5.0mL/MSVOA X/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T8-S-G2-(2.5)

Quant Time: Sep 05 14:51:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

