

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090419\
 Data File : VX012065.D
 Acq On : 04 Sep 2019 21:22
 Operator : SY/SP
 Sample : VSTDCCC050
 Misc : 5.00q/5.0mL/MSVOA X/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 05 05:52:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.461	0.439	4.8	87	0.00
3 P	Chloromethane	0.472	0.493	-4.4	93	0.00
4 C	Vinyl Chloride	0.444	0.452	-1.8#	89	0.00
5 T	Bromomethane	0.220	0.242	-10.0	98	0.00
6 T	Chloroethane	0.276	0.293	-6.2	94	0.00
7 T	Trichlorofluoromethane	0.690	0.690	0.0	88	0.00
8 T	Diethyl Ether	0.238	0.263	-10.5	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.511	0.502	1.8	86	0.00
10 T	Methyl Iodide	0.414	0.375	9.4	70	0.00
11 T	Tert butyl alcohol	0.066	0.068	-3.0	97	0.00
12 CM	1,1-Dichloroethene	0.492	0.503	-2.2#	89	0.00
13 T	Acrolein	0.021	0.012	42.9#	62	0.00
14 T	Allyl chloride	0.891	0.944	-5.9	92	0.00
15 T	Acrylonitrile	0.150	0.161	-7.3	96	0.00
16 T	Acetone	0.165	0.154	6.7	83	0.00
17 T	Carbon Disulfide	1.439	1.557	-8.2	87	0.00
18 T	Methyl Acetate	0.388	0.510	-31.4#	123	0.00
19 T	Methyl tert-butyl Ether	1.603	1.750	-9.2	97	0.00
20 T	Methylene Chloride	0.662	0.650	1.8	93	0.00
21 T	trans-1,2-Dichloroethene	0.578	0.612	-5.9	91	0.00
22 T	Diisopropyl ether	1.816	1.938	-6.7	95	0.00
23 T	Vinyl Acetate	1.175	1.215	-3.4	88	0.00
24 P	1,1-Dichloroethane	1.053	1.119	-6.3	94	0.00
25 T	2-Butanone	0.210	0.225	-7.1	93	0.00
26 T	2,2-Dichloropropane	0.972	0.881	9.4	81	0.00
27 T	cis-1,2-Dichloroethene	0.677	0.717	-5.9	93	0.00
28 T	Bromochloromethane	0.420	0.453	-7.9	91	0.00
29 T	Tetrahydrofuran	0.126	0.141	-11.9	99	0.00
30 C	Chloroform	1.130	1.200	-6.2#	95	0.00
31 T	Cyclohexane	0.846	0.822	2.8	86	0.00
32 T	1,1,1-Trichloroethane	0.984	1.030	-4.7	92	0.00
33 S	1,2-Dichloroethane-d4	0.690	0.816	-18.3	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.307	0.341	-11.1	99	0.00
36 T	1,1-Dichloropropene	0.440	0.426	3.2	88	0.00
37 T	Ethyl Acetate	0.239	0.254	-6.3	95	0.00
38 T	Carbon Tetrachloride	0.464	0.458	1.3	92	0.00
39 T	Methylcyclohexane	0.511	0.459	10.2	84	0.00
40 TM	Benzene	1.264	1.286	-1.7	93	0.00
41 T	Methacrylonitrile	0.143	0.157	-9.8	101	0.00
42 TM	1,2-Dichloroethane	0.455	0.484	-6.4	98	0.00
43 T	Isopropyl Acetate	0.487	0.513	-5.3	97	0.00
44 TM	Trichloroethene	0.341	0.340	0.3	92	0.00
45 C	1,2-Dichloropropane	0.327	0.327	0.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090419\
 Data File : VX012065.D
 Acq On : 04 Sep 2019 21:22
 Operator : SY/SP
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA X/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 05 05:52:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.191	0.199	-4.2	96	0.00
47 T	Bromodichloromethane	0.485	0.494	-1.9	93	0.00
48 T	Methyl methacrylate	0.245	0.261	-6.5	97	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	97	0.00
50 S	Toluene-d8	1.128	1.233	-9.3	94	0.00
51 T	4-Methyl-2-Pentanone	0.248	0.262	-5.6	98	0.00
52 CM	Toluene	0.823	0.821	0.2#	92	0.00
53 T	t-1,3-Dichloropropene	0.502	0.495	1.4	91	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.553	0.9	91	0.00
55 T	1,1,2-Trichloroethane	0.273	0.275	-0.7	93	0.00
56 T	Ethyl methacrylate	0.390	0.409	-4.9	96	0.00
57 T	1,3-Dichloropropane	0.470	0.482	-2.6	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.176	0.176	0.0	88	0.00
59 T	2-Hexanone	0.181	0.188	-3.9	97	0.00
60 T	Dibromochloromethane	0.325	0.333	-2.5	95	0.00
61 T	1,2-Dibromoethane	0.263	0.269	-2.3	94	0.00
62 S	4-Bromofluorobenzene	0.525	0.542	-3.2	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.312	0.341	-9.3	103	0.00
65 PM	Chlorobenzene	0.987	0.972	1.5	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.363	-0.6	96	0.00
67 C	Ethyl Benzene	1.798	1.760	2.1#	92	0.00
68 T	m/p-Xylenes	0.663	0.647	2.4	91	0.00
69 T	o-Xylene	0.642	0.634	1.2	93	0.00
70 T	Styrene	1.116	1.112	0.4	92	0.00
71 P	Bromoform	0.198	0.201	-1.5	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.669	3.610	1.6	91	0.00
74 T	N-amyl acetate	1.085	1.091	-0.6	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.765	0.775	-1.3	96	0.00
76 T	1,2,3-Trichloropropane	0.611	0.695	-13.7	109	0.00
77 T	Bromobenzene	0.820	0.834	-1.7	94	0.00
78 T	n-propylbenzene	4.432	4.229	4.6	89	0.00
79 T	2-Chlorotoluene	2.685	2.620	2.4	93	0.00
80 T	1,3,5-Trimethylbenzene	3.215	3.157	1.8	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.258	0.235	8.9	87	0.00
82 T	4-Chlorotoluene	3.200	3.166	1.1	91	0.00
83 T	tert-Butylbenzene	3.064	2.904	5.2	90	0.00
84 T	1,2,4-Trimethylbenzene	3.304	3.240	1.9	92	0.00
85 T	sec-Butylbenzene	3.806	3.617	5.0	89	0.00
86 T	p-Isopropyltoluene	3.461	3.334	3.7	91	0.00
87 T	1,3-Dichlorobenzene	1.644	1.602	2.6	93	0.00
88 T	1,4-Dichlorobenzene	1.644	1.605	2.4	93	0.00
89 T	n-Butylbenzene	3.362	3.104	7.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090419\
Data File : VX012065.D
Acq On : 04 Sep 2019 21:22
Operator : SY/SP
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA X/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 05 05:52:56 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.691	0.656	5.1	92	0.00
91 T	1,2-Dichlorobenzene	1.481	1.485	-0.3	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.160	0.164	-2.5	94	0.00
93 T	1,2,4-Trichlorobenzene	1.054	0.986	6.5	89	0.00
94 T	Hexachlorobutadiene	0.531	0.483	9.0	87	0.00
95 T	Naphthalene	2.281	2.291	-0.4	93	0.00
96 T	1,2,3-Trichlorobenzene	0.927	0.884	4.6	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5