

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090518\
 Data File : VX004367.D
 Acq On : 05 Sep 2018 13:08
 Operator : JC/MD
 Sample : J4693-03ME
 Misc : 5.78g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-MW-E(15-20)ME

Quant Time: Sep 06 01:17:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	262242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	415995	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	411595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	264948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	179675	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
35) Dibromofluoromethane	5.50	113	154113	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	8.71	98	595642	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	11.14	95	239061	54.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.70%	
Target Compounds						
20) Methylene Chloride	2.84	84	3326	1.01	ug/l	90
84) 1,2,4-Trimethylbenzene	11.81	105	12804	0.87	ug/l	97
89) n-Butylbenzene	12.39	91	4895	0.39	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090518\
Data File : VX004367.D
Acq On : 05 Sep 2018 13:08
Operator : JC/MD
Sample : J4693-03ME
Misc : 5.78g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-MW-E(15-20)ME

Quant Time: Sep 06 01:17:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

