

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090518\
 Data File : VX004371.D
 Acq On : 05 Sep 2018 15:31
 Operator : JC/MD
 Sample : J4794-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-180904

Quant Time: Sep 06 01:35:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	321447	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	504575	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	488298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	290425	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	211115	50.39	ug/l	0.00
Spiked Amount						
			Recovery	=	100.78%	
35) Dibromofluoromethane	5.50	113	181335	48.31	ug/l	0.00
Spiked Amount						
			Recovery	=	96.62%	
50) Toluene-d8	8.71	98	708156	49.98	ug/l	0.00
Spiked Amount						
			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.14	95	269207	50.92	ug/l	0.00
Spiked Amount						
			Recovery	=	101.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	2.86	84	2208	0.55	ug/l	95
78) n-propylbenzene	11.36	91	16008	0.73	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	74697	4.61	ug/l	99
85) sec-Butylbenzene	11.95	105	11023	0.58	ug/l	98
89) n-Butylbenzene	12.39	91	11870	0.86	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090518\
Data File : VX004371.D
Acq On : 05 Sep 2018 15:31
Operator : JC/MD
Sample : J4794-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-180904

Quant Time: Sep 06 01:35:28 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

