

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090518\
 Data File : VX004375.D
 Acq On : 05 Sep 2018 17:34
 Operator : JC/MD
 Sample : J4801-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OILY-WATER

Quant Time: Sep 06 12:48:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	300344	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	472569	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	443317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	234590	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	201806	51.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.10%	
35) Dibromofluoromethane	5.50	113	176034	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	
50) Toluene-d8	8.71	98	665658	50.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.32%	
62) 4-Bromofluorobenzene	11.14	95	239736	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090518\
Data File : VX004375.D
Acq On : 05 Sep 2018 17:34
Operator : JC/MD
Sample : J4801-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OILY-WATER

Quant Time: Sep 06 12:48:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

