

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX090518\
 Data File : VX004361.D
 Acq On : 05 Sep 2018 10:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

MMDadoda
 9/6/2018 12:48:21 PM

Quant Time: Sep 06 00:57:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	345052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	485185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	458429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	288485	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.06	65	202585	45.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.10%	
35) Dibromofluoromethane	5.50	113	179571	49.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.50%	
50) Toluene-d8	8.71	98	664795	48.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.58%	
62) 4-Bromofluorobenzene	11.14	95	259414	51.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.06%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.20	85	186184	46.59	ug/l	97
3) Chloromethane	1.32	50	200594	45.44	ug/l	98
4) Vinyl Chloride	1.40	62	212537	47.01	ug/l	98
5) Bromomethane	1.64	94	95064	42.35	ug/l	98
6) Chloroethane	1.71	64	139955	43.26	ug/l	98
7) Trichlorofluoromethane	1.92	101	295357	47.67	ug/l	95
8) Diethyl Ether	2.18	74	130644	49.24	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.38	101	190383	49.06	ug/l	99
10) Methyl Iodide	2.51	142	208811	47.88	ug/l	97
11) Tert butyl alcohol	3.06	59	205709	214.58	ug/l	99
12) 1,1-Dichloroethene	2.37	96	171676	46.97	ug/l	96
13) Acrolein	2.29	56	46616	201.70	ug/l	99
14) Allyl chloride	2.73	41	325559	48.47	ug/l	99
15) Acrylonitrile	3.15	53	541692	228.16	ug/l	99
16) Acetone	2.44	43	491794	268.48	ug/l	100
17) Carbon Disulfide	2.57	76	432037	42.68	ug/l	99
18) Methyl Acetate	2.78	43	383056	57.40	ug/l	100
19) Methyl tert-butyl Ether	3.20	73	605901	50.30	ug/l	98
20) Methylene Chloride	2.85	84	199944	46.18	ug/l	96
21) trans-1,2-Dichloroethene	3.16	96	184687	45.65	ug/l	98
22) Diisopropyl ether	3.87	45	636942	50.79	ug/l	99
23) Vinyl Acetate	3.82	43	2411331	237.26	ug/l	100
24) 1,1-Dichloroethane	3.70	63	356906	48.84	ug/l	99
25) 2-Butanone	4.70	43	737769	244.76	ug/l	99
26) 2,2-Dichloropropane	4.58	77	294251	51.34	ug/l	100
27) cis-1,2-Dichloroethene	4.59	96	222867	48.57	ug/l	100
28) Bromochloromethane	5.01	49	165838	51.26	ug/l	100
29) Tetrahydrofuran	5.15	42	446666	231.16	ug/l	100
30) Chloroform	5.21	83	358448	48.92	ug/l	98
31) Cyclohexane	5.57	56	307586	48.78	ug/l	98
32) 1,1,1-Trichloroethane	5.49	97	299983	49.10	ug/l	100
36) 1,1-Dichloropropene	5.79	75	269261	49.36	ug/l	99
37) Ethyl Acetate	4.85	43	255223	47.58	ug/l	99
38) Carbon Tetrachloride	5.77	117	260687	50.80	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090518\
 Data File : VX004361.D
 Acq On : 05 Sep 2018 10:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050

Manual Integrations
 APPROVED

MMDadoda
 9/6/2018 12:48:21 PM

Quant Time: Sep 06 00:57:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	332828	53.44	ug/l	98
40) Benzene	6.14	78	819820	50.30	ug/l	100
41) Methacrylonitrile	5.06	41	155856	50.12	ug/l	99
42) 1,2-Dichloroethane	6.20	62	273636	49.22	ug/l	99
43) Isopropyl Acetate	6.46	43	423382	51.36	ug/l	99
44) Trichloroethene	7.21	130	225662	50.02	ug/l	98
45) 1,2-Dichloropropane	7.51	63	211836	51.93	ug/l	98
46) Dibromomethane	7.66	93	134840	49.64	ug/l	100
47) Bromodichloromethane	7.90	83	266218	52.67	ug/l	99
48) Methyl methacrylate	7.77	41	228841	51.88	ug/l	98
49) 1,4-Dioxane	7.75	88	92188	920.57	ug/l	100
51) 4-Methyl-2-Pentanone	8.65	43	1407376	234.98	ug/l	99
52) Toluene	8.78	92	520275	51.00	ug/l	100
53) t-1,3-Dichloropropene	9.04	75	303342	54.34	ug/l	99
54) cis-1,3-Dichloropropene	8.43	75	334177	54.96	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	211540	51.36	ug/l	99
56) Ethyl methacrylate	9.18	69	311738	55.60	ug/l	98
57) 1,3-Dichloropropane	9.37	76	355703	51.24	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.31	63	664386	231.27	ug/l	100
59) 2-Hexanone	9.49	43	1088300	238.18	ug/l	100
60) Dibromochloromethane	9.59	129	215726	53.76	ug/l	100
61) 1,2-Dibromoethane	9.67	107	216966	50.91	ug/l	99
64) Tetrachloroethene	9.34	164	233251	51.51	ug/l	98
65) Chlorobenzene	10.14	112	569931	49.54	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.22	131	201658	50.82	ug/l	99
67) Ethyl Benzene	10.25	91	977900	52.34	ug/l	98
68) m/p-Xylenes	10.36	106	776212	106.24	ug/l	97
69) o-Xylene	10.70	106	362421	52.44	ug/l	99
70) Styrene	10.71	104	629433	56.25	ug/l	97
71) Bromoform	10.86	173	158491	49.55	ug/l	100
73) Isopropylbenzene	11.02	105	1040617	54.92	ug/l	100
74) N-amyl acetate	10.90	43	362432	49.68	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.27	83	305321	46.05	ug/l	99
76) 1,2,3-Trichloropropane	11.30	75	288463m	49.16	ug/l	
77) Bromobenzene	11.26	156	259067	48.74	ug/l	98
78) n-propylbenzene	11.36	91	1193996	54.79	ug/l	99
79) 2-Chlorotoluene	11.42	91	671433	50.06	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	831272	52.98	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	97401	53.86	ug/l	96
82) 4-Chlorotoluene	11.51	91	800120	50.69	ug/l	99
83) tert-Butylbenzene	11.77	119	854880	54.15	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	894469	55.53	ug/l	100
85) sec-Butylbenzene	11.94	105	1090095	58.16	ug/l	100
86) p-Isopropyltoluene	12.07	119	983983	58.58	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	518345	51.80	ug/l	99
88) 1,4-Dichlorobenzene	12.10	146	521915	49.47	ug/l	99
89) n-Butylbenzene	12.39	91	871340	63.79	ug/l	99
90) Hexachloroethane	12.60	117	139531	59.43	ug/l	99
91) 1,2-Dichlorobenzene	12.39	146	513137	56.11	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.00	75	59579	51.02	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090518\
Data File : VX004361.D
Acq On : 05 Sep 2018 10:09
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM

Quant Time: Sep 06 00:57:38 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	359568	54.03	ug/l	99
94) Hexachlorobutadiene	13.79	225	199290	64.42	ug/l	100
95) Naphthalene	13.83	128	1056999	56.16	ug/l	100
96) 1,2,3-Trichlorobenzene	14.02	180	394554	56.01	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.66 min Scan# 832

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampled :

VSTDCCC050

Tot Ion:168 Resp: 345052

Ion Ratio Lower Upper

168 100


99 49.3 39.9 59.9

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



5.66

100000

50000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 46.59 ug/l

RT: 1.20 min Scan# 100

Delta R.T. -0.00 min

Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion: 85 Resp: 186184

Ion Ratio Lower Upper

85 100

87 32.2 17.0 50.9



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.20

200000


150000


100000

50000

0

Time-->





#3

Chloromethane

Concen: 45.44 ug/l

RT: 1.32 min Scan# 121

Delta R.T. 0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 50 Resp: 200594

Ion Ratio Lower Upper

50 100


52 31.7 26.2 39.2

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

1.32



150000

100000

50000

0

Time-->



#4

Vinyl Chloride

Concen: 47.01 ug/l

RT: 1.40 min Scan# 134

Delta R.T. -0.00 min

Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion: 62 Resp: 212537

Ion Ratio Lower Upper

62 100

64 31.3 25.8 38.8



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0

1.40

200000


150000


100000

50000

0

Time-->





#5

Bromomethane

Concen: 42.35 ug/l

RT: 1.64 min Scan# 172

Delta R.T. 0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampled :

VSTDCCC050

Tot Ion: 94 Resp: 95064

Ion Ratio Lower Upper

94 100


96 94.7 74.5 111.7

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

100000

80000

60000



40000

20000

0

1.55 1.60 1.65 1.70

Time-->



#6

Chloroethane

Concen: 43.26 ug/l

RT: 1.71 min Scan# 185

Delta R.T. -0.00 min

Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion: 64 Resp: 139955

Ion Ratio Lower Upper

64 100

66 31.6 24.6 36.8



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

100000

80000

60000


40000


20000

0

1.60 1.70 1.80

Time-->





#7





Trichlorofluoromethane
 Concen: 47.67 ug/l
 RT: 1.92 min Scan# 219
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
101	100		
103	64.8	48.9	73.3

Manual Integrations
 APPROVED




MMDadoda
 9/6/2018 12:48:21 PM




#8

Diethyl Ether
 Concen: 49.24 ug/l
 RT: 2.18 min Scan# 262
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
74	100		
45	95.2	48.3	144.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 49.06 ug/l

RT: 2.38 min Scan# 294

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X





ClientSampled :

VSTDCCC050

Tot Ion:	101	Resp:	190383
Ion Ratio	Lower	Upper	
101	100		
85	42.1	34.2	51.4
151	81.5	65.0	97.6

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#10

Methyl Iodide

Concen: 47.88 ug/l




RT: 2.51 min Scan# 315


Delta R.T. 0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion:	142	Resp:	208811
Ion Ratio	Lower	Upper	
142	100		
127	40.3	33.8	50.6
141	13.8	11.4	17.0









#11
Tert butyl alcohol
Concen: 214.58 ug/l
RT: 3.06 min Scan# 406
Delta R.T. -0.02 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
59	100		
57	9.9	8.2	12.4




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#12
1,1-Dichloroethene
Concen: 46.97 ug/l
RT: 2.37 min Scan# 293
Delta R.T. 0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
96	100		
61	155.6	128.8	193.2
98	63.0	52.2	78.2









#13
 Acrolein
 Concen: 201.70 ug/l
 RT: 2.29 min Scan# 280
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	56	55	Ratio	100	69.5	Resp	46616	Lower	54.7	Upper	82.1
Ion											




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#14
 Allyl chloride
 Concen: 48.47 ug/l
 RT: 2.73 min Scan# 351
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	41	39	76	Ratio	100	59.9	33.9	Resp	325559	Lower	48.9	26.7	Upper	73.3	40.1
Ion															





#15

Acrylonitrile

Concen: 228.16 ug/l

RT: 3.15 min Scan# 420

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X





ClientSampled :

VSTDCCC050

Tot Ion:	53	Resp:	541692
Ion	Ratio	Lower	Upper
53	100		
52	81.9	65.2	97.8
51	34.7	28.2	42.2

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#16

Acetone

Concen: 268.48 ug/l




RT: 2.44 min Scan# 304


Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion:	43	Resp:	491794
Ion	Ratio	Lower	Upper
43	100		
58	33.1	26.2	39.4









#17
Carbon Disulfide
Concen: 42.68 ug/l
RT: 2.57 min Scan# 325
Delta R.T. 0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
76	100		
78	9.3	7.0	10.6




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#18
Methyl Acetate
Concen: 57.40 ug/l
RT: 2.78 min Scan# 359
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
43	100		
74	24.3	19.4	29.2





#19





Methyl tert-butyl Ether
 Concen: 50.30 ug/l
 RT: 3.20 min Scan# 428
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
73	100		
57	22.2	17.1	25.7

Manual Integrations
 APPROVED




MMDadoda
 9/6/2018 12:48:21 PM




#20

Methylene Chloride
 Concen: 46.18 ug/l
 RT: 2.85 min Scan# 372
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
84	100		
49	120.6	101.2	151.8
51	37.5	29.5	44.3
86	63.6	52.3	78.5





#21

trans-1,2-Dichloroethene

Concen: 45.65 ug/l

RT: 3.16 min Scan# 422

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 96 Resp: 184687

Ion Ratio Lower Upper

96 100

61 139.6 113.8 170.6


98 64.0 51.8 77.8

Manual Integrations




APPROVED

MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 95.90 (95.60 to 96.60): VX0
Ion 60.90 (60.60 to 61.60): VX0
Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 50.79 ug/l

RT: 3.87 min Scan# 538

Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion: 45 Resp: 636942


Ion Ratio Lower Upper

45 100



43 54.8 42.8 64.2


87 28.4 22.6 34.0

59 11.9 9.6 14.4



Abundance Ion 45.00 (44.70 to 45.70): VX0
Ion 43.00 (42.70 to 43.70): VX0
Ion 87.00 (86.70 to 87.70): VX0
Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 237.26 ug/l

RT: 3.82 min Scan# 530

Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tot Ion: 43 Resp: 2411331

Ion Ratio Lower Upper

43 100

86 11.2 8.9 13.3

Instrument :

MSVOA_X

ClientSampled :


VSTDCCC050

Manual Integrations

APPROVED




MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 48.84 ug/l

RT: 3.70 min Scan# 510

Delta R.T. 0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09


Tgt Ion: 63 Resp: 356906

Ion Ratio Lower Upper

63 100

98 7.4 3.5 10.4



100 4.8 2.4 7.1




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0





#25

2-Butanone

Concen: 244.76 ug/l

RT: 4.70 min Scan# 674

Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampled :

VSTDCCC050

Tot Ion: 43 Resp: 737769

Ion Ratio Lower Upper

43 100


72 27.0 22.0 33.0

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

300000

250000

200000

150000



100000

50000

0

4.60 4.70 4.80

Time-->



#26

2,2-Dichloropropane

Concen: 51.34 ug/l

RT: 4.58 min Scan# 655

Delta R.T. -0.00 min

Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion: 77 Resp: 294251

Ion Ratio Lower Upper

77 100

97 23.3 11.7 35.0



Abundance Ion 76.90 (76.60 to 77.60): VX0

Ion 96.90 (96.60 to 97.60): VX0

100000

80000

60000


40000


20000

0

4.40 4.50 4.60 4.70

Time-->





#27


cis-1,2-Dichloroethene
 Concen: 48.57 ug/l
 RT: 4.59 min Scan# 657
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050




Tot Ion	Ratio	Resp	Lower	Upper
96	100	222867		
61	141.4	0.0	282.6	
98	64.2	0.0	130.2	

Manual Integrations
 APPROVED

MMDadoda
 9/6/2018 12:48:21 PM




Abundance Ion 95.90 (95.60 to 96.60): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 97.90 (97.60 to 98.60): VX0





#28


Bromochloromethane
 Concen: 51.26 ug/l
 RT: 5.01 min Scan# 726
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Resp	Lower	Upper
49	100	165838		
129	2.1	0.0	4.2	
130	84.7	67.5	101.3	



Abundance Ion 48.90 (48.60 to 49.60): VX0
 Ion 128.80 (128.50 to 129.50): V
 Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 231.16 ug/l
RT: 5.15 min Scan# 749
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampled :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
42	100		
72	47.0	37.4	56.0
71	43.3	34.5	51.7




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#30
Chloroform
Concen: 48.92 ug/l
RT: 5.21 min Scan# 758
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
83	100		
85	64.1	52.6	79.0





#31

Cyclohexane

Concen: 48.78 ug/l

RT: 5.57 min Scan# 817

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	56	Resp:	307586
Ion	Ratio	Lower	Upper
56	100		
69	30.4	25.4	38.2
84	87.7	71.4	107.2

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#32

1,1,1-Trichloroethane

Concen: 49.10 ug/l




RT: 5.49 min Scan# 805


Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion:	97	Resp:	299983
Ion	Ratio	Lower	Upper
97	100		
99	64.8	52.0	78.0
61	43.5	34.9	52.3









#33
 1,2-Dichloroethane-d4
 Concen: 45.05 ug/l
 RT: 6.06 min Scan# 898
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
65	100	202585		
67	53.7	0.0	106.8	




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.86 min Scan# 1029
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Resp	Lower	Upper
114	100	485185		
63	19.1	0.0	39.0	
88	14.9	0.0	30.4	









#35
 Dibromofluoromethane
 Concen: 49.75 ug/l
 RT: 5.50 min Scan# 806
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
113	100	179571		
111	102.4	82.4	123.6	
192	24.3	19.4	29.2	




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#36
 1,1-Dichloropropene
 Concen: 49.36 ug/l
 RT: 5.79 min Scan# 854
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Resp	Lower	Upper
75	100	269261		
110	36.9	18.0	54.0	
77	31.4	24.6	36.8	





#37

Ethyl Acetate

Concen: 47.58 ug/l

RT: 4.85 min Scan# 700

Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	43	Resp:	255223
Ion	Ratio	Lower	Upper
43	100		
61	14.5	11.5	17.3
70	11.7	9.0	13.6

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#38

Carbon Tetrachloride

Concen: 50.80 ug/l




RT: 5.77 min Scan# 851


Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion:	117	Resp:	260687
Ion	Ratio	Lower	Upper
117	100		
119	94.6	78.8	118.2
121	30.6	24.9	37.3









#39
Methylvlcyclohexane
Concen: 53.44 ug/l
RT: 7.46 min Scan# 1127
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
83	100			
55	74.2	61.0	91.4	
98	48.3	39.6	59.4	




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#40
Benzene
Concen: 50.30 ug/l
RT: 6.14 min Scan# 911
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Resp	Lower	Upper
78	100			
77	23.6	19.0	28.4	









#41
Methacrylonitrile
Concen: 50.12 ug/l
RT: 5.06 min Scan# 733
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
41	100		
39	52.4	42.4	63.6
67	74.7	59.2	88.8
52	29.5	23.0	34.4




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#42
1,2-Dichloroethane
Concen: 49.22 ug/l
RT: 6.20 min Scan# 920
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
62	100		
98	10.6	0.0	20.6





#43

Isopropyl Acetate

Concen: 51.36 ug/l

RT: 6.46 min Scan# 963

Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X





ClientSampled :

VSTDCCC050

Tot Ion:	43	Resp:	423382
Ion	Ratio	Lower	Upper
43	100		
61	21.4	17.0	25.4
87	14.1	11.4	17.2

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#44

Trichloroethene

Concen: 50.02 ug/l




RT: 7.21 min Scan# 1086


Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion:	130	Resp:	225662
Ion	Ratio	Lower	Upper
130	100		
95	93.0	0.0	181.6









#45
 1,2-Dichloropropane
 Concen: 51.93 ug/l
 RT: 7.51 min Scan# 1136
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
63	100		
65	31.5	24.3	36.5




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#46
 Dibromomethane
 Concen: 49.64 ug/l
 RT: 7.66 min Scan# 1160
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
93	100		
95	82.2	65.5	98.3
174	118.0	94.2	141.4





#47

Bromodichloromethane

Concen: 52.67 ug/l

RT: 7.90 min Scan# 1199

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X


ClientSampleId :

VSTDCCC050




Tot Ion:	83	Resp:	266218
Ion Ratio	Lower	Upper	
83	100		
85	62.8	50.9	76.3
127	8.8	7.3	10.9

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



Abundance Ion 82.90 (82.60 to 83.60): VX0
Ion 84.90 (84.60 to 85.60): VX0
Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 51.88 ug/l


RT: 7.77 min Scan# 1179

Delta R.T. -0.00 min



Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion:	41	Resp:	228841
Ion Ratio	Lower	Upper	
41	100		
69	87.3	70.2	105.4
39	50.8	42.7	64.1



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 39.00 (38.70 to 39.70): VX0









#49
1,4-Dioxane
Concen: 920.57 ug/l
RT: 7.75 min Scan# 1175
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
88	100		
43	30.8	24.7	37.1
58	69.5	55.3	82.9




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#50
Toluene-d8
Concen: 48.79 ug/l
RT: 8.71 min Scan# 1333
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
98	100		
100	66.7	52.9	79.3





#51

4-Methyl-2-Pentanone

Concen: 234.98 ug/l

RT: 8.65 min Scan# 1322

Delta R.T. -0.01 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 1407376

Ion Ratio Lower Upper

43 100


58 41.8 33.1 49.7

Manual Integrations

APPROVED




MMDadoda

9/6/2018 12:48:21 PM



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 51.00 ug/l

RT: 8.78 min Scan# 1344

Delta R.T. -0.01 min

Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion: 92 Resp: 520275

Ion Ratio Lower Upper



92 100


91 170.1 136.4 204.6



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0









#53
t-1,3-Dichloropropene
Concen: 54.34 ug/l
RT: 9.04 min Scan# 1387
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 75 Resp: 303342
Ion Ratio Lower Upper
75 100
77 31.6 24.6 37.0




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#54
cis-1,3-Dichloropropene
Concen: 54.96 ug/l
RT: 8.43 min Scan# 1287
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion: 75 Resp: 334177
Ion Ratio Lower Upper
75 100
77 31.3 26.2 39.2
39 44.1 36.4 54.6









#55
 1,1,2-Trichloroethane
 Concen: 51.36 ug/l
 RT: 9.21 min Scan# 1415
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion:	97	Resp:	211540
Ion	Ratio	Lower	Upper
97	100		
83	84.3	66.4	99.6
85	54.8	43.3	64.9
99	62.5	49.0	73.4




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#56
 Ethyl methacrylate
 Concen: 55.60 ug/l
 RT: 9.18 min Scan# 1410
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion:	69	Resp:	311738
Ion	Ratio	Lower	Upper
69	100		
41	62.4	51.0	76.4
39	30.7	26.1	39.1









#57
 1,3-Dichloropropane
 Concen: 51.24 ug/l
 RT: 9.37 min Scan# 1441
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050

Tot Ion: 76 Resp: 355703
 Ion Ratio Lower Upper
 76 100
 78 32.0 25.5 38.3




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#58
 2-Chloroethyl Vinyl ether
 Concen: 231.27 ug/l
 RT: 8.31 min Scan# 1267
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion: 63 Resp: 664386
 Ion Ratio Lower Upper
 63 100
 106 29.8 23.9 35.9









#59
2-Hexanone
Concen: 238.18 ug/l
RT: 9.49 min Scan# 1461
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 43 Resp: 1088300
Ion Ratio Lower Upper
43 100
58 58.1 29.0 87.0




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#60
Dibromochloromethane
Concen: 53.76 ug/l
RT: 9.59 min Scan# 1476
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion: 129 Resp: 215726
Ion Ratio Lower Upper
129 100
127 77.4 38.5 115.5





#61

1,2-Dibromoethane

Concen: 50.91 ug/l

RT: 9.67 min Scan# 1490

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 107 Resp: 216966

Ion Ratio Lower Upper

107 100





109 94.3 76.0 114.0

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



#62

4-Bromofluorobenzene

Concen: 51.03 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09




Tgt Ion: 95 Resp: 259414


Ion Ratio Lower Upper

95 100

174 91.5 0.0 185.2

176 88.5 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
117	100		
82	49.0	47.8	71.6
119	31.6	29.3	43.9




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#64
Tetrachloroethene
Concen: 51.51 ug/l
RT: 9.34 min Scan# 1435
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
164	100		
166	131.4	102.8	154.2
129	89.5	69.4	104.0
131	84.1	67.9	101.9









#65
Chlorobenzene
Concen: 49.54 ug/l
RT: 10.14 min Scan# 1567
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion:112 Resp: 569931
Ion Ratio Lower Upper
112 100
114 33.1 27.8 41.8




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#66
1,1,1,2-Tetrachloroethane
Concen: 50.82 ug/l
RT: 10.22 min Scan# 1580
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion:131 Resp: 201658
Ion Ratio Lower Upper
131 100
133 97.2 48.1 144.4
119 64.6 32.9 98.6









#67
Ethyl Benzene
Concen: 52.34 ug/l
RT: 10.25 min Scan# 1585
Delta R.T. -0.01 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampled :
VSTDCCC050

Tot Ion: 91 Resp: 977900
Ion Ratio Lower Upper
91 100
106 31.2 25.9 38.9




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#68
m/p-Xylenes
Concen: 106.24 ug/l
RT: 10.36 min Scan# 1603
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion: 106 Resp: 776212
Ion Ratio Lower Upper
106 100
91 192.5 157.1 235.7









#69
o-Xylene
Concen: 52.44 ug/l
RT: 10.70 min Scan# 1659
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
106	100		
91	208.7	105.1	315.1




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#70
Styrene
Concen: 56.25 ug/l
RT: 10.71 min Scan# 1661
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
104	100		
78	44.8	37.4	56.0
103	51.9	43.8	65.6





#71

Bromoform

Concen: 49.55 ug/l

RT: 10.86 min Scan# 1685

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X


ClientSampleId :

VSTDCCC050

Tot Ion	Ratio	Lower	Upper
173	100		
175	48.2	24.1	72.3
254	0.2	0.2	0.2

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM






Abundance

Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l


RT: 12.08 min Scan# 1885

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
152	100		
115	78.7	39.0	117.0
150	175.4	0.0	351.0






Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 54.92 ug/l

RT: 11.02 min Scan# 1711

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion:105 Resp: 1040617

Ion Ratio Lower Upper

105 100





120 26.8 13.5 40.4

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



#74

N-ethyl acetate

Concen: 49.68 ug/l

RT: 10.90 min Scan# 1691

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion: 43 Resp: 362432




Ion Ratio Lower Upper


43 100

70 47.9 37.4 56.0

55 27.5 21.8 32.8

61 26.8 20.6 30.8





#75

1,1,2,2-Tetrachloroethane

Concen: 46.05 ug/l

RT: 11.27 min Scan# 1752

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050

Tot Ion:	83	Resp:	305321
Ion Ratio	Lower	Upper	
83	100		
131	10.8	5.2	15.6
85	65.4	32.3	96.8

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#76

1,2,3-Trichloropropane

Concen: 49.16 ug/l m




RT: 11.30 min Scan# 1757


Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Tgt Ion:	75	Resp:	288463
Ion Ratio	Lower	Upper	
75	100		
77	39.6	20.2	60.6





#77

Bromobenzene

Concen: 48.74 ug/l

RT: 11.26 min Scan# 1750

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X


ClientSampleId :

VSTDCCC050




Tot Ion:	156	Resp:	259067
Ion Ratio	Lower	Upper	
156	100		
77	146.9	71.5	214.3
158	98.1	48.9	146.8

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



Abundance Ion 155.80 (155.50 to 156.50): V
Ion 77.00 (76.70 to 77.70): VX0
Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 54.79 ug/l


RT: 11.36 min Scan# 1767

Delta R.T. -0.00 min



Lab File: VX004361.D


Acq: 05 Sep 2018 10:09

Tgt Ion:	91	Resp:	1193996
Ion Ratio	Lower	Upper	
91	100		
120	23.6	11.9	35.9



Abundance Ion 91.00 (90.70 to 91.70): VX0
Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 50.06 ug/l
 RT: 11.42 min Scan# 1777
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050

Tot Ion: 91 Resp: 671433
 Ion Ratio Lower Upper
 91 100
 126 34.5 17.9 53.7




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#80
 1,3,5-Trimethylbenzene
 Concen: 52.98 ug/l
 RT: 11.51 min Scan# 1791
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion: 105 Resp: 831272
 Ion Ratio Lower Upper
 105 100
 120 51.4 25.6 76.8









#81
trans-1,4-Dichloro-2-butene
Concen: 53.86 ug/l
RT: 11.07 min Scan# 1720
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
75	100		
53	94.9	80.1	120.1
89	45.3	37.2	55.8




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#82
4-Chlorotoluene
Concen: 50.69 ug/l
RT: 11.51 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
91	100		
126	31.5	15.5	46.5









#83
 tert-Butylbenzene
 Concen: 54.15 ug/l
 RT: 11.77 min Scan# 1834
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
119	100	854880		
91	56.0		27.0	81.0
134	24.3		11.7	35.1




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#84
 1,2,4-Trimethylbenzene
 Concen: 55.53 ug/l
 RT: 11.80 min Scan# 1840
 Delta R.T. -0.01 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	894469		
120	46.6		23.2	69.6





#85

sec-Butylbenzene

Concen: 58.16 ug/l

RT: 11.94 min Scan# 1863

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09

Instrument :

MSVOA_X

ClientSampled :

VSTDCCC050

Tot Ion:105 Resp: 1090095

Ion Ratio Lower Upper

105 100





134 21.0 10.5 31.5

Manual Integrations

APPROVED

MMDadoda

9/6/2018 12:48:21 PM



#86

p-Isopropyltoluene

Concen: 58.58 ug/l

RT: 12.07 min Scan# 1883

Delta R.T. -0.00 min

Lab File: VX004361.D

Acq: 05 Sep 2018 10:09




Tgt Ion:119 Resp: 983983


Ion Ratio Lower Upper

119 100

134 26.7 13.4 40.2

91 21.9 10.9 32.7









#87
 1,3-Dichlorobenzene
 Concen: 51.80 ug/l
 RT: 12.02 min Scan# 1876
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
146	100		
111	36.9	18.7	56.1
148	63.8	32.1	96.5




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#88
 1,4-Dichlorobenzene
 Concen: 49.47 ug/l
 RT: 12.10 min Scan# 1888
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
146	100		
111	36.6	18.1	54.1
148	64.3	32.0	96.0









#89
n-Butylbenzene
Concen: 63.79 ug/l
RT: 12.39 min Scan# 1936
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampled :
VSTDCCC050

Tot Ion	Ion	Ratio	Resp	Lower	Upper
91	100		871340		
92	54.8	27.3	81.9		
134	27.9	13.6	40.7		




Manual Integrations
APPROVED


MMDadoda
9/6/2018 12:48:21 PM



#90
Hexachloroethane
Concen: 59.43 ug/l
RT: 12.60 min Scan# 1970
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
117	100		139531		
201	94.6	46.9	140.6		









#91
 1,2-Dichlorobenzene
 Concen: 56.11 ug/l
 RT: 12.39 min Scan# 1936
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampled :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
146	100		
111	39.2	19.4	58.1
148	63.4	32.8	98.3




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 51.02 ug/l
 RT: 13.00 min Scan# 2036
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
75	100		
155	96.8	48.0	144.2
157	126.6	61.4	184.1









#93
 1,2,4-Trichlorobenzene
 Concen: 54.03 ug/l
 RT: 13.65 min Scan# 2142
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion:	180	Resp:	359568
Ion Ratio	Lower	Upper	
180	100		
182	95.4	48.4	145.0
145	28.8	14.3	42.9




Manual Integrations
 APPROVED


MMDadoda
 9/6/2018 12:48:21 PM



#94
 Hexachlorobutadiene
 Concen: 64.42 ug/l
 RT: 13.79 min Scan# 2165
 Delta R.T. -0.00 min
 Lab File: VX004361.D
 Acq: 05 Sep 2018 10:09

Tgt Ion:	225	Resp:	199290
Ion Ratio	Lower	Upper	
225	100		
223	60.3	30.1	90.5
227	61.7	30.8	92.4









#95
Naphthalene
Concen: 56.16 ug/l
RT: 13.83 min Scan# 2173
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.2	15.2
129	10.9	8.6	12.8

Manual Integrations
APPROVED

MMDadoda
9/6/2018 12:48:21 PM



#96
1,2,3-Trichlorobenzene
Concen: 56.01 ug/l
RT: 14.02 min Scan# 2204
Delta R.T. -0.00 min
Lab File: VX004361.D
Acq: 05 Sep 2018 10:09

Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.9	48.5	145.6
145	28.7	14.9	44.9

