


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090518\
Data File : VX004360.D
Acq On : 05 Sep 2018 09:21
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Title : SW846 8260
Last Update : Thu Aug 23 05:39:44 2018



AutoFind: Scans 1729, 1730, 1731; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	4078	PASS
75	95	30	60	48.4	11283	PASS
95	95	100	100	100.0	23332	PASS
96	95	5	9	6.5	1527	PASS
173	174	0.00	2	1.2	251	PASS
174	95	50	100	89.4	20866	PASS
175	174	5	9	7.7	1604	PASS
176	174	95	101	95.7	19970	PASS
177	176	5	9	6.8	1356	PASS