

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090519\
 Data File : VX012104.D
 Acq On : 05 Sep 2019 16:42
 Operator : SY/SP
 Sample : K4545-01RE
 Misc : 5.64g/5.0mL/MSVOA X/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T16-17-E2-(0)RE

Quant Time: Sep 06 06:43:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	129175	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	225347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	161710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	39404	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	89647	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
35) Dibromofluoromethane	5.48	113	68030	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	8.71	98	244395	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	11.14	95	72436	30.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.26%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.03	59	2723	15.906	ug/l #	85
16) Acetone	2.43	43	47743	112.067	ug/l	97
29) Tetrahydrofuran	5.12	42	5069	15.514	ug/l #	89
39) Methylcyclohexane	7.45	83	2750	1.194	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090519\
Data File : VX012104.D
Acq On : 05 Sep 2019 16:42
Operator : SY/SP
Sample : K4545-01RE
Misc : 5.64g/5.0mL/MSVOA X/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T16-17-E2-(0)RE

Quant Time: Sep 06 06:43:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

