

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090519\
 Data File : VX012106.D
 Acq On : 05 Sep 2019 17:38
 Operator : SY/SP
 Sample : K4545-05RE
 Misc : 6.43g/5.0mL/MSVOA X/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T16-17-E2-(2.25)RE

Quant Time: Sep 06 06:46:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	117606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	196850	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	123753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	22297	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	80597	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
35) Dibromofluoromethane	5.48	113	60702	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	8.71	98	203102	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
62) 4-Bromofluorobenzene	11.14	95	49731	24.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	48.14%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.03	59	4215	27.043	ug/l #	92
16) Acetone	2.43	43	9872	25.452	ug/l	98
20) Methylene Chloride	2.84	84	8415	5.403	ug/l	91
29) Tetrahydrofuran	5.14	42	2651	8.912	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090519\
Data File : VX012106.D
Acq On : 05 Sep 2019 17:38
Operator : SY/SP
Sample : K4545-05RE
Misc : 6.43g/5.0mL/MSVOA X/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T16-17-E2-(2.25)RE

Quant Time: Sep 06 06:46:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

