

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090519\
 Data File : VX012127.D
 Acq On : 06 Sep 2019 04:43
 Operator : SY/SP
 Sample : K4580-01
 Misc : 4.22g/5.0mL/MSVOA X/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T12-A3-(3.5)

Quant Time: Sep 06 07:33:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	155838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	272184	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	232851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97209	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	108061	50.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.50%	
35) Dibromofluoromethane	5.48	113	80657	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
50) Toluene-d8	8.71	98	319589	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.14	95	122663	42.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.88%	
Target Compounds						
20) Methylene Chloride	2.84	84	12004	5.816	ug/l	91
29) Tetrahydrofuran	5.13	42	3747	9.506	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090519\
Data File : VX012127.D
Acq On : 06 Sep 2019 04:43
Operator : SY/SP
Sample : K4580-01
Misc : 4.22g/5.0mL/MSVOA X/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T12-A3-(3.5)

Quant Time: Sep 06 07:33:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

