

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090519\
 Data File : VX012129.D
 Acq On : 06 Sep 2019 05:39
 Operator : SY/SP
 Sample : K4580-04
 Misc : 4.03g/5.0mL/MSVOA X/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T12-B1-(0)

Quant Time: Sep 06 07:38:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	154370	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	273574	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	236530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98328	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	107522	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	5.48	113	81539	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	319654	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
62) 4-Bromofluorobenzene	11.14	95	127708	44.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.96%	
Target Compounds						
16) Acetone	2.43	43	3995	7.847	ug/l	99
20) Methylene Chloride	2.84	84	12684	6.204	ug/l	90
29) Tetrahydrofuran	5.12	42	2916	7.468	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090519\
Data File : VX012129.D
Acq On : 06 Sep 2019 05:39
Operator : SY/SP
Sample : K4580-04
Misc : 4.03q/5.0mL/MSVOA X/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T12-B1-(0)

Quant Time: Sep 06 07:38:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

