

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090519\
 Data File : VX012133.D
 Acq On : 06 Sep 2019 07:31
 Operator : SY/SP
 Sample : K4580-08
 Misc : 3.79g/5.0mL/MSVOA X/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T12-B1-(3.5)

Quant Time: Sep 06 11:36:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	151054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	266739	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	235277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	100205	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	107125	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
35) Dibromofluoromethane	5.48	113	80961	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	8.71	98	316560	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	11.13	95	128094	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
Target Compounds						
16) Acetone	2.43	43	12330	24.750	ug/l	96
20) Methylene Chloride	2.84	84	13392	6.694	ug/l	99
29) Tetrahydrofuran	5.12	42	2995	7.839	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090519\
Data File : VX012133.D
Acq On : 06 Sep 2019 07:31
Operator : SY/SP
Sample : K4580-08
Misc : 3.79µ/5.0mL/MSVOA X/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T12-B1-(3.5)

Quant Time: Sep 06 11:36:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

