

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037507.D
 Acq On : 05 Sep 2023 13:06
 Operator : JC/MD
 Sample : 04269-04DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11A-13.5-083123DL

Quant Time: Sep 06 02:53:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222726	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79021	40.500	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.000%
35) Dibromofluoromethane	5.385	113	68694	46.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.140%
50) Toluene-d8	8.647	98	259958	48.585	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	94006	44.707	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.420%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1244	0.831	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.319	101	1943	1.358	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	43231	23.155	ug/l	78
44) Trichloroethene	7.123	130	55399	32.506	ug/l	98
64) Tetrachloroethene	9.269	164	344	0.236	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037507.D
Acq On : 05 Sep 2023 13:06
Operator : JC/MD
Sample : 04269-04DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11A-13.5-083123DL

Quant Time: Sep 06 02:53:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

