

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037509.D
 Acq On : 05 Sep 2023 13:55
 Operator : JC/MD
 Sample : 04252-01DL 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-06-9.1-083023DL

Quant Time: Sep 06 02:54:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78109	40.792	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.580%
35) Dibromofluoromethane	5.385	113	67339	46.504	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.653	98	255854	48.710	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.420%
62) 4-Bromofluorobenzene	11.079	95	90063	43.631	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.260%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	20234	14.412	ug/l	94
20) Methylene Chloride	2.788	84	889	0.531	ug/l #	78
27) cis-1,2-Dichloroethene	4.489	96	2312	1.262	ug/l	77
44) Trichloroethene	7.129	130	107760	64.410	ug/l	96
64) Tetrachloroethene	9.275	164	361	0.254	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037509.D
Acq On : 05 Sep 2023 13:55
Operator : JC/MD
Sample : 04252-01DL 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-06-9.1-083023DL

Quant Time: Sep 06 02:54:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

