

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037510.D
 Acq On : 05 Sep 2023 14:19
 Operator : JC/MD
 Sample : 04266-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-04B-26.5-083123

Quant Time: Sep 06 02:54:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	128929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214269	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83497	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74785	40.234	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.460%
35) Dibromofluoromethane	5.385	113	64213	45.251	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.500%
50) Toluene-d8	8.647	98	248722	48.320	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.640%
62) 4-Bromofluorobenzene	11.079	95	89855	44.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.840%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.135	130	1879	1.146	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037510.D
Acq On : 05 Sep 2023 14:19
Operator : JC/MD
Sample : 04266-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-04B-26.5-083123

Quant Time: Sep 06 02:54:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

