

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037514.D
 Acq On : 05 Sep 2023 15:58
 Operator : JC/MD
 Sample : 04276-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 06 02:56:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171819	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68703	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71886	41.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.520%
35) Dibromofluoromethane	5.391	113	61763	46.983	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.960%
50) Toluene-d8	8.653	98	230154	48.265	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.540%
62) 4-Bromofluorobenzene	11.079	95	76700	40.929	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.860%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037514.D
Acq On : 05 Sep 2023 15:58
Operator : JC/MD
Sample : 04276-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 06 02:56:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

