

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037517.D
 Acq On : 05 Sep 2023 17:12
 Operator : JC/MD
 Sample : 04254-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20230829

Quant Time: Sep 06 02:57:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213262	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76241	40.943	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.880%
35) Dibromofluoromethane	5.385	113	65208	46.169	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.340%
50) Toluene-d8	8.653	98	249737	48.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.079	95	88351	43.882	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.760%
Target Compounds						
16) Acetone	2.386	43	2432	3.338	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037517.D
Acq On : 05 Sep 2023 17:12
Operator : JC/MD
Sample : 04254-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20230829

Quant Time: Sep 06 02:57:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

