

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037518.D
 Acq On : 05 Sep 2023 17:36
 Operator : JC/MD
 Sample : 04229-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01-9.5-082923RE

Quant Time: Sep 06 02:57:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206713	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73515	41.532	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.060%
35) Dibromofluoromethane	5.385	113	63844	46.635	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	8.647	98	241987	48.729	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.460%
62) 4-Bromofluorobenzene	11.079	95	84104	43.096	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.200%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3145	4.542	ug/l #	84
27) cis-1,2-Dichloroethene	4.489	96	46121	27.229	ug/l	79
44) Trichloroethene	7.129	130	160426	101.424	ug/l	95
64) Tetrachloroethene	9.275	164	497	0.370	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037518.D
Acq On : 05 Sep 2023 17:36
Operator : JC/MD
Sample : 04229-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01-9.5-082923RE

Quant Time: Sep 06 02:57:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

