

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037519.D
 Acq On : 05 Sep 2023 18:01
 Operator : JC/MD
 Sample : 04254-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW11-MW01I-20230830

Quant Time: Sep 06 02:57:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	125593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75575	41.739	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.480%
35) Dibromofluoromethane	5.391	113	63024	45.694	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.380%
50) Toluene-d8	8.647	98	243189	48.608	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	85836	43.657	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.320%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2104	2.970	ug/l #	67
44) Trichloroethene	7.135	130	1716	1.077	ug/l	84
52) Toluene	8.726	92	1980	0.533	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037519.D
Acq On : 05 Sep 2023 18:01
Operator : JC/MD
Sample : 04254-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW11-MW01I-20230830

Quant Time: Sep 06 02:57:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

