

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037529.D
 Acq On : 06 Sep 2023 03:36
 Operator : JC/MD
 Sample : VX0905WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0905WBL02

Quant Time: Sep 06 05:26:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241966	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86163	41.202	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.400%
35) Dibromofluoromethane	5.391	113	74411	46.435	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	284272	48.904	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.800%
62) 4-Bromofluorobenzene	11.079	95	102008	44.655	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037529.D
Acq On : 06 Sep 2023 03:36
Operator : JC/MD
Sample : VX0905WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0905WBL02

Quant Time: Sep 06 05:26:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

