

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037533.D
 Acq On : 06 Sep 2023 05:34
 Operator : JC/MD
 Sample : 04232-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SBG-4DL

Quant Time: Sep 06 06:27:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85320	41.361	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	82.720%		
35) Dibromofluoromethane	5.385	113	73301	46.776	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.560%		
50) Toluene-d8	8.647	98	282496	49.697	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.400%		
62) 4-Bromofluorobenzene	11.079	95	101862	45.599	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.200%		
Target Compounds						
16) Acetone	2.380	43	10786	13.366	ug/l #	89
20) Methylene Chloride	2.788	84	2868	1.589	ug/l	94
40) Benzene	6.038	78	202035	31.641	ug/l	97
43) Isopropyl Acetate	6.342	43	20353	5.637	ug/l #	92
52) Toluene	8.720	92	106660	25.282	ug/l	100
67) Ethyl Benzene	10.195	91	12176	1.525	ug/l	98
68) m/p-Xylenes	10.299	106	23998	7.785	ug/l	98
69) o-Xylene	10.640	106	14971	4.922	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	12024	2.019	ug/l	85
95) Naphthalene	13.774	128	72653	8.963	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037533.D
Acq On : 06 Sep 2023 05:34
Operator : JC/MD
Sample : 04232-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SBG-4DL

Quant Time: Sep 06 06:27:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

