

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037535.D
 Acq On : 06 Sep 2023 06:22
 Operator : JC/MD
 Sample : 04232-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SBG-6DL

Quant Time: Sep 06 07:45:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	141159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237224	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84143	41.347	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	82.700%		
35) Dibromofluoromethane	5.385	113	72946	46.431	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.860%		
50) Toluene-d8	8.647	98	277519	48.697	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.400%		
62) 4-Bromofluorobenzene	11.079	95	103006	45.993	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.980%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11690	14.683	ug/l	91
20) Methylene Chloride	2.788	84	2851	1.601	ug/l #	82
40) Benzene	6.038	78	342519	53.505	ug/l	98
43) Isopropyl Acetate	6.348	43	19868	5.489	ug/l #	90
52) Toluene	8.720	92	159405	37.687	ug/l	99
67) Ethyl Benzene	10.195	91	19421	2.443	ug/l	98
68) m/p-Xylenes	10.299	106	36689	11.953	ug/l	99
69) o-Xylene	10.640	106	22436	7.409	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	17600	2.939	ug/l	89
95) Naphthalene	13.774	128	135698	16.645	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037535.D
Acq On : 06 Sep 2023 06:22
Operator : JC/MD
Sample : 04232-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SBG-6DL

Quant Time: Sep 06 07:45:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

