

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037537.D
 Acq On : 06 Sep 2023 07:11
 Operator : JC/MD
 Sample : 04232-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SBG-8DL

Quant Time: Sep 06 07:46:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227060	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81090	40.597	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.200%
35) Dibromofluoromethane	5.391	113	69546	46.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.647	98	266495	48.856	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.079	95	94937	44.288	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.580%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	8882	11.367	ug/l	95
40) Benzene	6.037	78	293989	47.980	ug/l	98
43) Isopropyl Acetate	6.348	43	9695	2.798	ug/l #	90
52) Toluene	8.720	92	115880	28.623	ug/l	99
67) Ethyl Benzene	10.195	91	11334	1.509	ug/l	98
68) m/p-Xylenes	10.299	106	22542	7.774	ug/l	97
69) o-Xylene	10.640	106	13187	4.609	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	10342	1.913	ug/l	85
95) Naphthalene	13.774	128	79690	10.830	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037537.D
Acq On : 06 Sep 2023 07:11
Operator : JC/MD
Sample : 04232-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SBG-8DL

Quant Time: Sep 06 07:46:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

