

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
 Data File : VX037540.D
 Acq On : 06 Sep 2023 08:24
 Operator : JC/MD
 Sample : 04285-03DL 8000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 832DL

Quant Time: Sep 06 08:58:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140960	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	233346	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80557	39.640	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.280%
35) Dibromofluoromethane	5.385	113	70906	45.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.760%
50) Toluene-d8	8.647	98	273294	48.753	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.079	95	96572	43.837	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.680%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.965	59	5421	11.372	ug/l	# 96
40) Benzene	6.037	78	159901	25.393	ug/l	98
52) Toluene	8.720	92	36295	8.724	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090523\
Data File : VX037540.D
Acq On : 06 Sep 2023 08:24
Operator : JC/MD
Sample : 04285-03DL 8000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
832DL

Quant Time: Sep 06 08:58:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

