



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX090523\  
Data File : VX037500.D  
Acq On : 05 Sep 2023 09:24  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X081723W.M  
Title : SW846 8260  
Last Update : Fri Aug 18 09:30:38 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 15.7  | 5708  | PASS      |
| 75     | 95      | 30     | 60     | 48.5  | 17642 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 36411 | PASS      |
| 96     | 95      | 5      | 9      | 7.0   | 2545  | PASS      |
| 173    | 174     | 0.00   | 2      | 0.6   | 148   | PASS      |
| 174    | 95      | 50     | 100    | 70.6  | 25701 | PASS      |
| 175    | 174     | 5      | 9      | 7.6   | 1941  | PASS      |
| 176    | 174     | 95     | 101    | 97.6  | 25072 | PASS      |
| 177    | 176     | 5      | 9      | 6.6   | 1654  | PASS      |