

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090618\
 Data File : VX004394.D
 Acq On : 06 Sep 2018 19:20
 Operator : JC/MD
 Sample : J4781-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BS-STONES

Quant Time: Sep 07 07:47:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	291520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	471249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	452118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	247329	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	202295	53.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.48%	
35) Dibromofluoromethane	5.50	113	180025	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
50) Toluene-d8	8.72	98	672693	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	11.14	95	236199	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
Target Compounds						
16) Acetone	2.45	43	62122	33.04	ug/l	100
18) Methyl Acetate	2.78	43	13637	2.42	ug/l	97
20) Methylene Chloride	2.85	84	62928	17.20	ug/l	100
43) Isopropyl Acetate	6.46	43	213821	26.71	ug/l	100
95) Naphthalene	13.83	128	98372	6.10	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090618\
Data File : VX004394.D
Acq On : 06 Sep 2018 19:20
Operator : JC/MD
Sample : J4781-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BS-STONES

Quant Time: Sep 07 07:47:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

