

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090619\
 Data File : VX012149.D
 Acq On : 06 Sep 2019 15:42
 Operator : SY/SP
 Sample : K4601-06
 Misc : 5.15µ/5.0mL/MSVOA X/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T12-E1-(2.7)

Quant Time: Sep 07 01:33:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	151207	50.00	µg/l	0.00
34) 1,4-Difluorobenzene	6.85	114	265504	50.00	µg/l	0.00
63) Chlorobenzene-d5	10.11	117	223813	50.00	µg/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	88753	50.00	µg/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	105461	50.55	µg/l	0.00
Spiked Amount	50.000		Recovery	=	101.10%	
35) Dibromofluoromethane	5.48	113	78677	48.19	µg/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
50) Toluene-d8	8.71	98	308042	51.41	µg/l	0.00
Spiked Amount	50.000		Recovery	=	102.82%	
62) 4-Bromofluorobenzene	11.13	95	119370	42.84	µg/l	0.00
Spiked Amount	50.000		Recovery	=	85.68%	
Target Compounds						
20) Methylene Chloride	2.84	84	10860	5.423	µg/l	94
29) Tetrahydrofuran	5.12	42	3592	9.392	µg/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090619\
 Data File : VX012149.D
 Acq On : 06 Sep 2019 15:42
 Operator : SY/SP
 Sample : K4601-06
 Misc : 5.15µ/5.0mL/MSVOA X/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 T12-E1-(2.7)

Quant Time: Sep 07 01:33:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

