

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090619\
 Data File : VX012162.D
 Acq On : 06 Sep 2019 21:44
 Operator : SY/SP
 Sample : K4601-20
 Misc : 2.64g/5.0mL/MSVOA X/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T12-G1-(0)

Quant Time: Sep 07 03:12:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	155227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	269902	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	226264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	91006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	107774	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
35) Dibromofluoromethane	5.48	113	81668	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
50) Toluene-d8	8.71	98	317194	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
62) 4-Bromofluorobenzene	11.13	95	120441	42.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.04%	
Target Compounds						
16) Acetone	2.42	43	26538	51.838	ug/l	98
29) Tetrahydrofuran	5.13	42	3345	8.520	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090619\
Data File : VX012162.D
Acq On : 06 Sep 2019 21:44
Operator : SY/SP
Sample : K4601-20
Misc : 2.64g/5.0mL/MSVOA X/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T12-G1-(0)

Quant Time: Sep 07 03:12:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

