

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031127.D
 Acq On : 06 Sep 2022 14:40
 Operator : JC/MD
 Sample : VX0906WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01

Quant Time: Sep 06 17:29:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	167643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	261078	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111403	51.646	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.300%
35) Dibromofluoromethane	5.391	113	102453	53.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.540%
50) Toluene-d8	8.653	98	352047	50.206	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.420%
62) 4-Bromofluorobenzene	11.079	95	139372	56.275	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.540%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	30003	17.393	ug/l	96
3) Chloromethane	1.288	50	22016	14.431	ug/l	94
4) Vinyl Chloride	1.374	62	27449	15.233	ug/l	92
5) Bromomethane	1.611	94	16200	20.110	ug/l	94
6) Chloroethane	1.685	64	14307	12.861	ug/l	98
7) Trichlorofluoromethane	1.886	101	60951	20.018	ug/l	93
8) Diethyl Ether	2.136	74	18189	17.129	ug/l	78
9) 1,1,2-Trichlorotrifluo...	2.325	101	37668	18.677	ug/l #	88
10) Methyl Iodide	2.447	142	47451	25.805	ug/l	96
11) Tert butyl alcohol	2.995	59	51062	104.654	ug/l #	90
12) 1,1-Dichloroethene	2.319	96	36013	18.491	ug/l	88
13) Acrolein	2.239	56	38125	96.586	ug/l	99
14) Allyl chloride	2.666	41	41940	16.215	ug/l #	83
15) Acrylonitrile	3.069	53	95519	94.450	ug/l	94
16) Acetone	2.392	43	82484	99.053	ug/l	99
17) Carbon Disulfide	2.508	76	68398	13.523	ug/l	97
18) Methyl Acetate	2.709	43	45918	19.189	ug/l #	85
19) Methyl tert-butyl Ether	3.117	73	120551	20.099	ug/l	94
20) Methylene Chloride	2.788	84	39678	18.188	ug/l #	83
21) trans-1,2-Dichloroethene	3.093	96	36629	16.902	ug/l	94
22) Diisopropyl ether	3.770	45	90356	17.337	ug/l #	79
23) Vinyl Acetate	3.727	43	383211	91.034	ug/l #	91
24) 1,1-Dichloroethane	3.611	63	63326	18.033	ug/l	95
25) 2-Butanone	4.568	43	122107	94.982	ug/l #	86
26) 2,2-Dichloropropane	4.483	77	56902	20.305	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	45707	18.818	ug/l	91
28) Bromochloromethane	4.904	49	17147	16.481	ug/l #	42
29) Tetrahydrofuran	5.013	42	73383	88.541	ug/l #	73
30) Chloroform	5.099	83	74562	19.362	ug/l	92
31) Cyclohexane	5.477	56	46483	15.775	ug/l #	83
32) 1,1,1-Trichloroethane	5.385	97	68578	20.774	ug/l	95
36) 1,1-Dichloropropene	5.696	75	48182	18.054	ug/l	96
37) Ethyl Acetate	4.715	43	44496	19.190	ug/l #	91
38) Carbon Tetrachloride	5.684	117	59069	21.324	ug/l	95
39) Methylcyclohexane	7.385	83	59043	17.852	ug/l	93
40) Benzene	6.038	78	151543	19.779	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031127.D
 Acq On : 06 Sep 2022 14:40
 Operator : JC/MD
 Sample : VX0906WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01

Quant Time: Sep 06 17:29:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	24325	19.677	ug/l #	85
42) 1,2-Dichloroethane	6.092	62	56628	22.405	ug/l	98
43) Isopropyl Acetate	6.348	43	79531	21.946	ug/l #	94
44) Trichloroethene	7.129	130	51149	21.091	ug/l	84
45) 1,2-Dichloropropane	7.434	63	37751	19.222	ug/l #	88
46) Dibromomethane	7.586	93	31056	21.128	ug/l #	89
47) Bromodichloromethane	7.824	83	59056	21.505	ug/l	96
48) Methyl methacrylate	7.696	41	36056	19.708	ug/l #	82
49) 1,4-Dioxane	7.684	88	22134	397.757	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	232046	99.606	ug/l	88
52) Toluene	8.720	92	99743	20.074	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	60551	22.269	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	63967	21.039	ug/l #	90
55) 1,1,2-Trichloroethane	9.153	97	44490	21.061	ug/l	94
56) Ethyl methacrylate	9.116	69	61977	21.461	ug/l #	85
57) 1,3-Dichloropropane	9.311	76	72365	21.666	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	134306	93.135	ug/l #	88
59) 2-Hexanone	9.433	43	185148	105.095	ug/l	89
60) Dibromochloromethane	9.525	129	50302	22.933	ug/l	98
61) 1,2-Dibromoethane	9.610	107	49745	22.306	ug/l	100
64) Tetrachloroethene	9.275	164	44607	18.562	ug/l	97
65) Chlorobenzene	10.079	112	113905	19.815	ug/l	96
66) 1,1,1,2-Tetrachloroethane	10.165	131	46691	21.997	ug/l	97
67) Ethyl Benzene	10.195	91	195901	19.892	ug/l	98
68) m/p-Xylenes	10.305	106	155246	40.194	ug/l	100
69) o-Xylene	10.646	106	78255	20.658	ug/l	94
70) Styrene	10.659	104	129221	21.079	ug/l	99
71) Bromoform	10.799	173	36813	22.305	ug/l #	97
73) Isopropylbenzene	10.963	105	205254	20.266	ug/l	100
74) N-amyl acetate	10.848	43	58717	19.207	ug/l #	81
75) 1,1,2,2-Tetrachloroethane	11.213	83	63606	19.639	ug/l	96
76) 1,2,3-Trichloropropane	11.238	75	76431	25.734	ug/l	86
77) Bromobenzene	11.201	156	54127	20.728	ug/l	85
78) n-propylbenzene	11.305	91	231026	20.028	ug/l	97
79) 2-Chlorotoluene	11.366	91	142376	20.044	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	180270	21.151	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	20090	21.258	ug/l	96
82) 4-Chlorotoluene	11.457	91	175473	22.078	ug/l	99
83) tert-Butylbenzene	11.719	119	191095	22.527	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	203923	24.072	ug/l	94
85) sec-Butylbenzene	11.890	105	214724	19.325	ug/l	98
86) p-Isopropyltoluene	12.006	119	190016	21.594	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	101282	20.249	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	106054	21.290	ug/l	96
89) n-Butylbenzene	12.335	91	161337	20.334	ug/l	97
90) Hexachloroethane	12.542	117	35981	22.563	ug/l	91
91) 1,2-Dichlorobenzene	12.335	146	117704	22.944	ug/l	94
92) 1,2-Dibromo-3-Chloropr...	12.945	75	18955	24.020	ug/l	86
93) 1,2,4-Trichlorobenzene	13.591	180	74230	22.863	ug/l	97
94) Hexachlorobutadiene	13.725	225	29070	20.449	ug/l	95
95) Naphthalene	13.780	128	259889	22.821	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	76065	22.334	ug/l	92


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031127.D
Acq On : 06 Sep 2022 14:40
Operator : JC/MD
Sample : VX0906WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01

Quant Time: Sep 06 17:29:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:168 Resp: 167643
Ion Ratio Lower Upper
168 100
99 53.7 44.0 66.0



#2
Dichlorodifluoromethane
Concen: 17.393 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 85 Resp: 30003
Ion Ratio Lower Upper
85 100
87 35.0 16.5 49.5









#3
Chloromethane
Concen: 14.431 ug/l
RT: 1.288 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 50 Resp: 22016
Ion Ratio Lower Upper
50 100
52 36.3 26.3 39.5



#4
Vinyl Chloride
Concen: 15.233 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 62 Resp: 27449
Ion Ratio Lower Upper
62 100
64 36.1 25.2 37.8









#5
Bromomethane
Concen: 20.110 ug/l
RT: 1.611 min Scan# 86
Delta R.T. 0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 94 Resp: 16200
Ion Ratio Lower Upper
94 100
96 90.0 76.9 115.3



#6
Chloroethane
Concen: 12.861 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 64 Resp: 14307
Ion Ratio Lower Upper
64 100
66 34.9 27.2 40.8









#7
Trichlorofluoromethane
Concen: 20.018 ug/l
RT: 1.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:101 Resp: 60951
Ion Ratio Lower Upper
101 100
103 67.3 49.7 74.5



#8
Diethyl Ether
Concen: 17.129 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 74 Resp: 18189
Ion Ratio Lower Upper
74 100
45 80.6 51.5 154.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.677 ug/l
RT: 2.325 min Scan# 20
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:101 Resp: 37668
Ion Ratio Lower Upper
101 100
85 42.1 34.0 51.0
151 89.2 59.4 89.0#



#10
Methyl Iodide
Concen: 25.805 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion:142 Resp: 47451
Ion Ratio Lower Upper
142 100
127 41.2 30.2 45.2
141 13.9 11.3 16.9









#11
Tert butyl alcohol
Concen: 104.654 ug/l
RT: 2.995 min Scan# 313
Delta R.T. 0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 59 Resp: 51062
Ion Ratio Lower Upper
59 100
57 5.7 7.4 11.2#



#12
1,1-Dichloroethene
Concen: 18.491 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 96 Resp: 36013
Ion Ratio Lower Upper
96 100
61 127.8 117.4 176.0
98 61.0 52.2 78.4









#13
Acrolein
Concen: 96.586 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 56 Resp: 38125
Ion Ratio Lower Upper
56 100
55 71.1 56.6 84.8



#14
Allyl chloride
Concen: 16.215 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 41 Resp: 41940
Ion Ratio Lower Upper
41 100
39 70.8 48.9 73.3
76 47.0 26.9 40.3#









#15
Acrylonitrile
Concen: 94.450 ug/l
RT: 3.069 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 53 Resp: 95519
Ion Ratio Lower Upper
53 100
52 86.9 65.4 98.0
51 37.4 27.7 41.5



#16
Acetone
Concen: 99.053 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 43 Resp: 82484
Ion Ratio Lower Upper
43 100
58 35.3 27.7 41.5









#17
Carbon Disulfide
Concen: 13.523 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 76 Resp: 68398
Ion Ratio Lower Upper
76 100
78 8.5 7.8 11.6



#18
Methyl Acetate
Concen: 19.189 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 43 Resp: 45918
Ion Ratio Lower Upper
43 100
74 32.0 19.4 29.2#





#19

Methyl tert-butyl Ether

Concen: 20.099 ug/l

RT: 3.117 min Scan# 311

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 73 Resp: 120551

Ion Ratio Lower Upper

73 100

57 20.9 19.0 28.6



#20

Methylene Chloride

Concen: 18.188 ug/l

RT: 2.788 min Scan# 279

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Tgt Ion: 84 Resp: 39678




Ion Ratio Lower Upper


84 100

49 90.2 93.8 140.6#

51 27.5 29.4 44.0#

86 65.6 51.0 76.4









#21
trans-1,2-Dichloroethene
Concen: 16.902 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 96 Resp: 36629
Ion Ratio Lower Upper
96 100
61 124.3 105.6 158.4
98 66.0 50.1 75.1



#22
Diisopropyl ether
Concen: 17.337 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 45 Resp: 90356
Ion Ratio Lower Upper
45 100
43 62.5 65.4 98.0#
87 40.6 22.3 33.5#
59 14.2 9.0 13.4#









#23
 Vinyl Acetate
 Concen: 91.034 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. -0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01


Tgt Ion: 43 Resp: 383211
 Ion Ratio Lower Upper
 43 100
 86 14.2 8.5 12.7#



#24
 1,1-Dichloroethane
 Concen: 18.033 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40

Tgt Ion: 63 Resp: 63326
 Ion Ratio Lower Upper
 63 100
 98 8.4 3.5 10.5
 100 6.5 2.3 6.9





#25

2-Butanone

Concen: 94.982 ug/l

RT: 4.568 min Scan# 571

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 43 Resp: 122107

Ion Ratio Lower Upper

43 100

72 34.4 21.7 32.5#



#26

2,2-Dichloropropane

Concen: 20.305 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.006 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 77 Resp: 56902

Ion Ratio Lower Upper

77 100

97 23.2 13.0 39.0









#27
 cis-1,2-Dichloroethene
 Concen: 18.818 ug/l
 RT: 4.489 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01


Tgt Ion: 96 Resp: 45707
 Ion Ratio Lower Upper
 96 100
 61 120.1 0.0 270.0
 98 65.9 0.0 127.8



#28
 Bromochloromethane
 Concen: 16.481 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40

Tgt Ion: 49 Resp: 17147
 Ion Ratio Lower Upper
 49 100
 129 3.2 0.0 4.4
 130 130.2 62.8 94.2#





#29

Tetrahydrofuran

Concen: 88.541 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion: 42 Resp: 73383

Ion Ratio Lower Upper

42 100

72 64.6 36.5 54.7#

71 58.8 33.9 50.9#



#30

Chloroform

Concen: 19.362 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 83 Resp: 74562

Ion Ratio Lower Upper

83 100

85 71.1 51.5 77.3









#31
Cyclohexane
Concen: 15.775 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 56 Resp: 46483
Ion Ratio Lower Upper
56 100
69 35.2 25.4 38.0
84 105.8 68.9 103.3#



#32
1,1,1-Trichloroethane
Concen: 20.774 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 97 Resp: 68578
Ion Ratio Lower Upper
97 100
99 64.1 51.4 77.0
61 36.5 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 51.646 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 65 Resp: 111403

Ion Ratio Lower Upper

65 100

67 51.8 0.0 109.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion: 114 Resp: 261078


Ion Ratio Lower Upper

114 100

63 16.1 0.0 38.4

88 15.2 0.0 31.8





#35

Dibromofluoromethane

Concen: 53.774 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion:113 Resp: 102453

Ion Ratio Lower Upper

113 100

111 100.8 83.2 124.8

192 20.3 14.8 22.2



#36

1,1-Dichloropropene

Concen: 18.054 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion: 75 Resp: 48182


Ion Ratio Lower Upper

75 100

110 42.0 19.3 57.9

77 32.0 24.4 36.6





#37

Ethyl Acetate

Concen: 19.190 ug/l

RT: 4.715 min Scan# 596

Delta R.T. -0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :





MSVOA_X

ClientSampleId :

VX0906WBS01

Tgt Ion: 43 Resp: 44496

Ion	Ratio	Lower	Upper
43	100		
61	18.0	11.7	17.5#
70	15.7	9.3	13.9#



#38

Carbon Tetrachloride

Concen: 21.324 ug/l

RT: 5.684 min Scan# 754




Delta R.T. -0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Tgt Ion: 117 Resp: 59069

Ion	Ratio	Lower	Upper
117	100		
119	87.2	74.1	111.1
121	30.0	25.8	38.8





#39

Methylcyclohexane

Concen: 17.852 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion: 83 Resp: 59043

Ion Ratio Lower Upper

83 100

55 66.9 57.7 86.5

98 51.4 36.5 54.7



#40

Benzene

Concen: 19.779 ug/l

RT: 6.038 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 78 Resp: 151543

Ion Ratio Lower Upper

78 100

77 24.7 18.0 27.0









#41
Methacrylonitrile
Concen: 19.677 ug/l
RT: 4.934 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:	41	Resp:	24325
Ion	Ratio	Lower	Upper
41	100		
39	56.5	42.9	64.3
67	94.5	60.1	90.1#
52	37.3	22.6	34.0#



#42
1,2-Dichloroethane
Concen: 22.405 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion:	62	Resp:	56628
Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	20.2





#43

Isopropyl Acetate

Concen: 21.946 ug/l

RT: 6.348 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion: 43 Resp: 79531

Ion Ratio Lower Upper

43 100

61 22.2 16.5 24.7

87 17.6 11.1 16.7#



#44

Trichloroethene

Concen: 21.091 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion:130 Resp: 51149

Ion Ratio Lower Upper

130 100

95 81.5 0.0 194.8





#45

1,2-Dichloropropane

Concen: 19.222 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 63 Resp: 37751

Ion Ratio Lower Upper

63 100

65 38.1 25.2 37.8#



#46

Dibromomethane

Concen: 21.128 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion: 93 Resp: 31056


Ion Ratio Lower Upper

93 100

95 84.3 67.4 101.0

174 115.0 75.6 113.4#









#47
Bromodichloromethane
Concen: 21.505 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 83 Resp: 59056
Ion Ratio Lower Upper
83 100
85 64.0 53.5 80.3
127 7.4 6.6 10.0



#48
Methyl methacrylate
Concen: 19.708 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 41 Resp: 36056
Ion Ratio Lower Upper
41 100
69 110.1 68.8 103.2#
39 60.0 44.6 66.8





#49

1,4-Dioxane

Concen: 397.757 ug/l

RT: 7.684 min Scan# 11

Delta R.T. 0.007 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion: 88 Resp: 22134

Ion Ratio Lower Upper

88 100

43 25.8 23.8 35.8

58 63.7 53.8 80.6



#50

Toluene-d8

Concen: 50.206 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 98 Resp: 352047

Ion Ratio Lower Upper

98 100

100 64.7 52.3 78.5









#51
4-Methyl-2-Pentanone
Concen: 99.606 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 43 Resp: 232046
Ion Ratio Lower Upper
43 100
58 46.6 31.4 47.2



#52
Toluene
Concen: 20.074 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 92 Resp: 99743
Ion Ratio Lower Upper
92 100
91 175.4 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 22.269 ug/l

RT: 8.982 min Scan# 1195

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 75 Resp: 60551

Ion Ratio Lower Upper

75 100

77 33.9 26.3 39.5



#54

cis-1,3-Dichloropropene

Concen: 21.039 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion: 75 Resp: 63967


Ion Ratio Lower Upper

75 100

77 30.8 25.0 37.4

39 35.9 37.5 56.3#





#55

1,1,2-Trichloroethane

Concen: 21.061 ug/l

RT: 9.153 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01

Tgt Ion: 97 Resp: 44490





Ion Ratio Lower Upper

97 100

83 79.9 68.9 103.3

85 57.1 44.0 66.0

99 67.5 48.3 72.5



#56

Ethyl methacrylate

Concen: 21.461 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion: 69 Resp: 61977


Ion Ratio Lower Upper

69 100

41 50.9 52.3 78.5#

39 29.8 28.2 42.4





#57

1,3-Dichloropropane

Concen: 21.666 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 76 Resp: 72365

Ion Ratio Lower Upper

76 100

78 32.0 25.9 38.9



#58

2-Chloroethyl Vinyl ether

Concen: 93.135 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.013 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 63 Resp: 134306

Ion Ratio Lower Upper

63 100

106 34.0 22.2 33.2#









#59
2-Hexanone
Concen: 105.095 ug/l
RT: 9.433 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 43 Resp: 185148
Ion Ratio Lower Upper
43 100
58 63.1 27.6 82.8



#60
Dibromochloromethane
Concen: 22.933 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 129 Resp: 50302
Ion Ratio Lower Upper
129 100
127 75.2 38.5 115.5









#61
1,2-Dibromoethane
Concen: 22.306 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:107 Resp: 49745
Ion Ratio Lower Upper
107 100
109 96.2 76.6 114.8



#62
4-Bromofluorobenzene
Concen: 56.275 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 95 Resp: 139372
Ion Ratio Lower Upper
95 100
174 83.0 0.0 149.0
176 80.6 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:117 Resp: 240689
Ion Ratio Lower Upper
117 100
82 49.3 42.7 64.1
119 32.4 25.4 38.2



#64
Tetrachloroethene
Concen: 18.562 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion:164 Resp: 44607
Ion Ratio Lower Upper
164 100
166 125.1 98.2 147.4
129 94.7 80.8 121.2
131 94.0 75.5 113.3









#65
Chlorobenzene
Concen: 19.815 ug/l
RT: 10.079 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:112 Resp: 113905
Ion Ratio Lower Upper
112 100
114 32.7 24.6 36.8



#66
1,1,1,2-Tetrachloroethane
Concen: 21.997 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion:131 Resp: 46691
Ion Ratio Lower Upper
131 100
133 95.5 46.9 140.8
119 64.8 34.4 103.1









#67
Ethyl Benzene
Concen: 19.892 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 91 Resp: 195901
Ion Ratio Lower Upper
91 100
106 30.8 25.5 38.3



#68
m/p-Xylenes
Concen: 40.194 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 106 Resp: 155246
Ion Ratio Lower Upper
106 100
91 198.1 158.2 237.4





#69

o-Xylene

Concen: 20.658 ug/l

RT: 10.646 min Scan# 1568

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion:106 Resp: 78255

Ion Ratio Lower Upper

106 100

91 216.8 103.8 311.6



#70

Styrene

Concen: 21.079 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. -0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion:104 Resp: 129221


Ion Ratio Lower Upper

104 100

78 48.6 37.8 56.8

103 55.6 44.5 66.7





#71

Bromoform

Concen: 22.305 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion:173 Resp: 36813

Ion Ratio Lower Upper

173 100

175 49.6 23.8 71.2

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion:152 Resp: 125641


Ion Ratio Lower Upper

152 100

115 69.9 40.5 121.5

150 167.4 0.0 347.0









#73
Isopropylbenzene
Concen: 20.266 ug/l
RT: 10.963 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:105 Resp: 205254
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4



#74
N-ethyl acetate
Concen: 19.207 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 43 Resp: 58717
Ion Ratio Lower Upper
43 100
70 56.1 32.3 48.5#
55 30.1 18.4 27.6#
61 30.9 19.2 28.8#









#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.639 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01


Tgt Ion: 83 Resp: 63606
 Ion Ratio Lower Upper
 83 100
 131 12.2 4.5 13.7
 85 67.7 32.5 97.5



#76
 1,2,3-Trichloropropane
 Concen: 25.734 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40

Tgt Ion: 75 Resp: 76431
 Ion Ratio Lower Upper
 75 100
 77 31.2 19.8 59.3









#77
Bromobenzene
Concen: 20.728 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:156 Resp: 54127
Ion Ratio Lower Upper
156 100
77 134.1 82.7 247.9
158 95.1 47.5 142.5



#78
n-propylbenzene
Concen: 20.028 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 91 Resp: 231026
Ion Ratio Lower Upper
91 100
120 24.9 11.6 34.8





#79

2-Chlorotoluene

Concen: 20.044 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :





VX0906WBS01

Tgt Ion: 91 Resp: 142376

Ion Ratio Lower Upper

91 100

126 38.0 17.4 52.3



#80

1,3,5-Trimethylbenzene

Concen: 21.151 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.006 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 105 Resp: 180270

Ion Ratio Lower Upper

105 100

120 50.2 24.4 73.2









#81
trans-1,4-Dichloro-2-butene
Concen: 21.258 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion: 75 Resp: 20090
Ion Ratio Lower Upper
75 100
53 91.7 74.2 111.4
89 49.2 34.8 52.2



#82
4-Chlorotoluene
Concen: 22.078 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 91 Resp: 175473
Ion Ratio Lower Upper
91 100
126 29.8 15.3 45.9





#83

tert-Butylbenzene

Concen: 22.527 ug/l

RT: 11.719 min Scan# 1744

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion:119 Resp: 191095

Ion Ratio Lower Upper

119 100

91 53.3 28.3 84.9

134 22.1 11.2 33.5



#84

1,2,4-Trimethylbenzene

Concen: 24.072 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. 0.000 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion:105 Resp: 203923

Ion Ratio Lower Upper

105 100

120 41.7 22.9 68.8









#85
 sec-Butylbenzene
 Concen: 19.325 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.006 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01


Tgt Ion:105 Resp: 214724
 Ion Ratio Lower Upper
 105 100
 134 21.1 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 21.594 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.006 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40

Tgt Ion:119 Resp: 190016
 Ion Ratio Lower Upper
 119 100
 134 26.5 13.9 41.6
 91 22.0 11.8 35.3









#87
 1,3-Dichlorobenzene
 Concen: 20.249 ug/l
 RT: 11.969 min Scan# 1785
 Delta R.T. 0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906WBS01


Tgt Ion:146 Resp: 101282
 Ion Ratio Lower Upper
 146 100
 111 37.9 20.3 60.8
 148 62.6 31.8 95.3



#88
 1,4-Dichlorobenzene
 Concen: 21.290 ug/l
 RT: 12.043 min Scan# 1797
 Delta R.T. -0.000 min
 Lab File: VX031127.D
 Acq: 06 Sep 2022 14:40

Tgt Ion:146 Resp: 106054
 Ion Ratio Lower Upper
 146 100
 111 37.9 20.3 60.9
 148 67.9 32.3 96.8





#89

n-Butylbenzene

Concen: 20.334 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion: 91 Resp: 161337

Ion Ratio Lower Upper

91 100

92 53.5 26.2 78.5

134 28.4 12.6 37.8



#90

Hexachloroethane

Concen: 22.563 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX031127.D




Acq: 06 Sep 2022 14:40


Tgt Ion: 117 Resp: 35981

Ion Ratio Lower Upper

117 100

201 76.0 34.2 102.5









#91
1,2-Dichlorobenzene
Concen: 22.944 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40




Instrument :
MSVOA_X
ClientSampleId :
VX0906WBS01


Tgt Ion:146 Resp: 117704
Ion Ratio Lower Upper
146 100
111 36.4 20.7 62.1
148 60.1 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 24.020 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX031127.D
Acq: 06 Sep 2022 14:40

Tgt Ion: 75 Resp: 18955
Ion Ratio Lower Upper
75 100
155 101.1 45.0 134.8
157 130.1 56.9 170.7





#93

1,2,4-Trichlorobenzene

Concen: 22.863 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion:180 Resp: 74230

Ion Ratio Lower Upper

180 100

182 94.0 48.2 144.6

145 30.1 15.9 47.6



#94

Hexachlorobutadiene

Concen: 20.449 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40




Tgt Ion:225 Resp: 29070


Ion Ratio Lower Upper

225 100

223 67.5 30.6 91.6

227 65.0 31.9 95.9





#95

Naphthalene

Concen: 22.821 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. 0.006 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Instrument :

MSVOA_X

ClientSampleId :

VX0906WBS01





Tgt Ion:128 Resp: 259889

Ion Ratio Lower Upper

128 100

127 13.2 10.2 15.2

129 11.3 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 22.334 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX031127.D

Acq: 06 Sep 2022 14:40

Tgt Ion:180 Resp: 76065

Ion Ratio Lower Upper

180 100

182 85.4 47.0 141.0

145 30.3 17.0 50.9

