

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031125.D
 Acq On : 06 Sep 2022 13:53
 Operator : JC/MD
 Sample : VX0906MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0906MBL01

Quant Time: Sep 06 17:28:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	196922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	307797	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140181	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	119673	47.231	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.460%
35) Dibromofluoromethane	5.385	113	114465	50.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.920%
50) Toluene-d8	8.647	98	402308	48.665	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.340%
62) 4-Bromofluorobenzene	11.079	95	141768	48.554	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031125.D
Acq On : 06 Sep 2022 13:53
Operator : JC/MD
Sample : VX0906MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0906MBL01

Quant Time: Sep 06 17:28:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

