

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031134.D
 Acq On : 06 Sep 2022 17:36
 Operator : JC/MD
 Sample : N4438-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KTS2-WC-S-MH-01-0-20

Quant Time: Sep 06 18:24:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	230509	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	346543	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134593	45.379	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.760%
35) Dibromofluoromethane	5.391	113	127842	50.551	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.100%
50) Toluene-d8	8.653	98	462401	49.681	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.085	95	165919	50.472	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	23750	20.742	ug/l	95
20) Methylene Chloride	2.788	84	10220	1.545	ug/l	# 83
25) 2-Butanone	4.581	43	7278	4.117	ug/l	# 84
43) Isopropyl Acetate	6.342	43	95659	19.887	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031134.D
Acq On : 06 Sep 2022 17:36
Operator : JC/MD
Sample : N4438-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KTS2-WC-S-MH-01-0-20

Quant Time: Sep 06 18:24:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

