

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031135.D
 Acq On : 06 Sep 2022 17:59
 Operator : JC/MD
 Sample : N4438-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KTS2-WC-S-MH-02-0-20

Quant Time: Sep 06 18:25:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	360500	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151257	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	142113	48.334	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.660%
35) Dibromofluoromethane	5.391	113	129955	49.397	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.653	98	462431	47.760	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	162009	47.374	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.740%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	24125	21.254	ug/l	97
20) Methylene Chloride	2.794	84	18009	4.528	ug/l #	77
37) Ethyl Acetate	4.727	43	16072m	5.020	ug/l	
43) Isopropyl Acetate	6.348	43	89027	17.791	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031135.D
Acq On : 06 Sep 2022 17:59
Operator : JC/MD
Sample : N4438-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KTS2-WC-S-MH-02-0-20

Quant Time: Sep 06 18:25:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

