

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031136.D
 Acq On : 06 Sep 2022 18:23
 Operator : JC/MD
 Sample : N4489-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ORA-0045DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/07/2022
 Supervised By : Mahesh Dadoda 09/07/2022

Quant Time: Sep 06 19:11:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	354212	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	308667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138710	48.016	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.040%		
35) Dibromofluoromethane	5.391	113	126273	48.850	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.700%		
50) Toluene-d8	8.653	98	460090	48.362	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.720%		
62) 4-Bromofluorobenzene	11.079	95	158814	47.264	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.520%		
Target Compounds						
16) Acetone	2.386	43	115976	103.993	ug/l	99
25) 2-Butanone	4.562	43	51953	30.175	ug/l #	85
37) Ethyl Acetate	4.721	43	27299m	8.678	ug/l	
59) 2-Hexanone	9.433	43	7142	2.988	ug/l	88
68) m/p-Xylenes	10.305	106	1966	0.397	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	3584	0.399	ug/l #	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031136.D
Acq On : 06 Sep 2022 18:23
Operator : JC/MD
Sample : N4489-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ORA-0045DL

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/07/2022
Supervised By : Mahesh Dadoda 09/07/2022

Quant Time: Sep 06 19:11:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

