

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031143.D
 Acq On : 06 Sep 2022 21:07
 Operator : JC/MD
 Sample : N4479-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-13

Quant Time: Sep 07 01:15:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	206453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	359681	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	292747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139145	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127220	47.892	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.780%
35) Dibromofluoromethane	5.391	113	121876	46.432	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.653	98	426502	44.150	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.300%
62) 4-Bromofluorobenzene	11.079	95	153639	45.029	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.060%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	19651	19.162	ug/l	92
20) Methylene Chloride	2.788	84	14668	3.856	ug/l #	76
37) Ethyl Acetate	4.714	43	22650	7.090	ug/l #	92
43) Isopropyl Acetate	6.342	43	102509	20.532	ug/l #	91
95) Naphthalene	13.780	128	14610	1.158	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031143.D
Acq On : 06 Sep 2022 21:07
Operator : JC/MD
Sample : N4479-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-13

Quant Time: Sep 07 01:15:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

