

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031145.D
 Acq On : 06 Sep 2022 21:54
 Operator : JC/MD
 Sample : N4478-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-SV28486

Quant Time: Sep 07 01:15:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	222363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	368047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	300287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	131541	45.975	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.960%
35) Dibromofluoromethane	5.391	113	133023	49.527	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	8.653	98	438348	44.345	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.680%
62) 4-Bromofluorobenzene	11.079	95	154576	44.274	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
16) Acetone	2.386	43	2417	2.188	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031145.D
Acq On : 06 Sep 2022 21:54
Operator : JC/MD
Sample : N4478-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-SV28486

Quant Time: Sep 07 01:15:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

