

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
 Data File : VX031146.D
 Acq On : 06 Sep 2022 22:17
 Operator : JC/MD
 Sample : PB147457TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147457TB

Quant Time: Sep 07 01:15:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	213876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	335443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	302878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135599	49.274	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.540%
35) Dibromofluoromethane	5.391	113	126690	51.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	8.653	98	447820	49.706	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.085	95	162665	51.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.240%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18735	17.635	ug/l	95
20) Methylene Chloride	2.794	84	11636	2.416	ug/l #	73
37) Ethyl Acetate	4.727	43	21449	7.200	ug/l #	91
43) Isopropyl Acetate	6.342	43	97663	20.975	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090622\
Data File : VX031146.D
Acq On : 06 Sep 2022 22:17
Operator : JC/MD
Sample : PB147457TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147457TB

Quant Time: Sep 07 01:15:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

