

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
 Data File : VX037560.D
 Acq On : 06 Sep 2023 17:28
 Operator : JC/MD
 Sample : 04287-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-090123

Quant Time: Sep 06 22:47:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	202568	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71680	40.594	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	81.180%		
35) Dibromofluoromethane	5.385	113	61818	46.079	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.160%		
50) Toluene-d8	8.653	98	236786	48.658	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.320%		
62) 4-Bromofluorobenzene	11.079	95	78778	41.193	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	82.380%#		
Target Compounds						
16) Acetone	2.380	43	1359	1.967	ug/l	93
20) Methylene Chloride	2.788	84	725	0.469	ug/l #	79
44) Trichloroethene	7.129	130	3816	2.462	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
Data File : VX037560.D
Acq On : 06 Sep 2023 17:28
Operator : JC/MD
Sample : 04287-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-090123

Quant Time: Sep 06 22:47:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

