

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
 Data File : VX037561.D
 Acq On : 06 Sep 2023 17:53
 Operator : JC/MD
 Sample : 04287-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10B-42.5-090123-3V

Quant Time: Sep 06 22:47:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203898	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72369	40.552	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	81.100%		
35) Dibromofluoromethane	5.391	113	63057	46.696	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.400%		
50) Toluene-d8	8.653	98	242588	49.525	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.040%		
62) 4-Bromofluorobenzene	11.079	95	81993	42.595	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.180%		
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	490	0.358	ug/l #	73
12) 1,1-Dichloroethene	2.319	96	27262	21.394	ug/l	80
20) Methylene Chloride	2.782	84	530	0.339	ug/l #	81
21) trans-1,2-Dichloroethene	3.093	96	2151	1.485	ug/l	95
24) 1,1-Dichloroethane	3.611	63	7594	2.992	ug/l #	88
27) cis-1,2-Dichloroethene	4.495	96	55440	32.465	ug/l	77
42) 1,2-Dichloroethane	6.092	62	5945	2.755	ug/l	91
44) Trichloroethene	7.129	130	50619	32.444	ug/l	92
49) 1,4-Dioxane	7.678	88	1533	37.514	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
Data File : VX037561.D
Acq On : 06 Sep 2023 17:53
Operator : JC/MD
Sample : 04287-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10B-42.5-090123-3V

Quant Time: Sep 06 22:47:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

