

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
 Data File : VX037563.D
 Acq On : 06 Sep 2023 18:42
 Operator : JC/MD
 Sample : 04287-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-02-133-090123-3V

Quant Time: Sep 06 22:48:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73980	39.799	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.600%
35) Dibromofluoromethane	5.385	113	65236	46.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.653	98	247097	48.244	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.480%
62) 4-Bromofluorobenzene	11.079	95	84421	41.942	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.880%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2652	3.647	ug/l	96
25) 2-Butanone	4.586	43	4070	3.717	ug/l	90
27) cis-1,2-Dichloroethene	4.495	96	600	0.337	ug/l	80
44) Trichloroethene	7.135	130	3072	1.883	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
Data File : VX037563.D
Acq On : 06 Sep 2023 18:42
Operator : JC/MD
Sample : 04287-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-02-133-090123-3V

Quant Time: Sep 06 22:48:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

