

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
 Data File : VX037566.D
 Acq On : 06 Sep 2023 19:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 06 22:49:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222850	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77911	39.820	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	79.640%		
35) Dibromofluoromethane	5.385	113	68063	46.117	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.240%		
50) Toluene-d8	8.647	98	260802	48.715	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.440%		
62) 4-Bromofluorobenzene	11.079	95	90463	42.998	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.000%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	4882	10.637	ug/l #	78
16) Acetone	2.386	43	750	0.980	ug/l	98
20) Methylene Chloride	2.788	84	494	0.289	ug/l #	86
25) 2-Butanone	4.580	43	312	0.271	ug/l #	51
27) cis-1,2-Dichloroethene	4.489	96	391	0.209	ug/l	91
31) Cyclohexane	5.556	56	2431	1.105	ug/l #	1
36) 1,1-Dichloropropene	5.550	75	9796	4.672	ug/l #	51
38) Carbon Tetrachloride	5.550	117	13881	6.094	ug/l #	15
44) Trichloroethene	7.129	130	2051	1.203	ug/l	90
51) 4-Methyl-2-Pentanone	8.647	43	967	0.468	ug/l #	1
76) 1,2,3-Trichloropropane	11.079	75	43230	24.155	ug/l #	39
81) trans-1,4-Dichloro-2-b...	11.079	75	43230	67.467	ug/l #	5

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
Data File : VX037566.D
Acq On : 06 Sep 2023 19:56
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 06 22:49:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

