

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
 Data File : VX037567.D
 Acq On : 06 Sep 2023 20:21
 Operator : JC/MD
 Sample : 04308-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-06B-76.5-090523

Quant Time: Sep 06 22:49:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74992	40.103	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.200%
35) Dibromofluoromethane	5.385	113	65304	45.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.700%
50) Toluene-d8	8.647	98	252870	48.943	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.880%
62) 4-Bromofluorobenzene	11.079	95	89981	44.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.640%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.313	96	35959	26.931	ug/l	82
21) trans-1,2-Dichloroethene	3.087	96	7953	5.240	ug/l	83
24) 1,1-Dichloroethane	3.611	63	5494	2.066	ug/l #	92
27) cis-1,2-Dichloroethene	4.489	96	92819	51.872	ug/l	78
40) Benzene	6.044	78	1768	0.305	ug/l	95
42) 1,2-Dichloroethane	6.092	62	3588	1.576	ug/l	100
44) Trichloroethene	7.129	130	14160	8.604	ug/l	93
49) 1,4-Dioxane	7.677	88	1631	37.839	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090623\
Data File : VX037567.D
Acq On : 06 Sep 2023 20:21
Operator : JC/MD
Sample : 04308-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-06B-76.5-090523

Quant Time: Sep 06 22:49:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

