

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090718\
 Data File : VX004404.D
 Acq On : 07 Sep 2018 13:44
 Operator : JC/MD
 Sample : VX0907MBL01
 Misc : 5.00µl/10mL/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0907MBL01

Quant Time: Sep 08 03:38:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	315038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	505982	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	463394	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	262861	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	215455	52.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.94%	
35) Dibromofluoromethane	5.50	113	189084	50.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.46%	
50) Toluene-d8	8.71	98	705899	49.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.36%	
62) 4-Bromofluorobenzene	11.14	95	247350	46.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090718\
Data File : VX004404.D
Acq On : 07 Sep 2018 13:44
Operator : JC/MD
Sample : VX0907MBL01
Misc : 5.0g/10mL/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0907MBL01

Quant Time: Sep 08 03:38:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

