

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090718\
 Data File : VX004410.D
 Acq On : 07 Sep 2018 16:31
 Operator : JC/MD
 Sample : J4812-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12

Quant Time: Sep 08 04:08:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	322235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	515032	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	500996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	276665	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	222870	53.07	ug/l	0.00
Spiked Amount			Recovery	=	106.14%	
35) Dibromofluoromethane	5.50	113	196143	51.19	ug/l	0.00
Spiked Amount			Recovery	=	102.38%	
50) Toluene-d8	8.72	98	734309	50.77	ug/l	0.00
Spiked Amount			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	11.14	95	261250	48.41	ug/l	0.00
Spiked Amount			Recovery	=	96.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
7) Trichlorofluoromethane	1.93	101	2745	0.47	ug/l	92
19) Methyl tert-butyl Ether	3.20	73	10951	0.97	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	30827	7.19	ug/l	90
30) Chloroform	5.21	83	5235	0.77	ug/l	86
44) Trichloroethene	7.21	130	222627	46.49	ug/l	100
52) Toluene	8.79	92	3303	0.30	ug/l	93
64) Tetrachloroethene	9.34	164	70719	14.29	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090718\
Data File : VX004410.D
Acq On : 07 Sep 2018 16:31
Operator : JC/MD
Sample : J4812-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: Sep 08 04:08:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

